

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100819\
 Data File : BE100616.D
 Acq On : 8 Oct 2019 19:43
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 09 01:46:28 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 09 01:42:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.674	0.647	4.0	104	0.00
3	n-Nitrosodimethylamine	0.317	0.344	-8.5	137	-0.01
4 S	2-Fluorophenol	1.080	1.150	-6.5	119	0.00
5 S	Phenol-d6	1.357	1.494	-10.1	126	0.00
6	bis(2-Chloroethyl)ether	1.257	1.237	1.6	110	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.250	0.239	4.4	120	0.00
9	Naphthalene	4.126	4.059	1.6	112	0.00
10	Hexachlorobutadiene	0.133	0.127	4.5	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.599	0.593	1.0	114	-0.01
12	2-Methylnaphthalene	0.678	0.673	0.7	116	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
14 S	2,4,6-Tribromophenol	0.091	0.092	-1.1	131	0.00
15 S	2-Fluorobiphenyl	1.451	1.411	2.8	115	0.00
16	Acenaphthylene	1.699	1.620	4.6	119	0.00
17	Acenaphthene	1.138	1.094	3.9	117	0.00
18	Fluorene	1.454	1.415	2.7	121	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.198	0.183	7.6	116	0.00
21	Hexachlorobenzene	0.167	0.160	4.2	114	0.00
22	Pentachlorophenol	0.046	0.041	10.9	147	-0.01
23	Phenanthrene	1.138	1.136	0.2	117	0.00
24	Anthracene	0.979	0.914	6.6	120	0.00
25 SURR	Fluoranthene-d10	4.934	4.677	5.2	120	0.00
26	Fluoranthene	1.401	1.369	2.3	120	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
28	Pvrene	1.025	1.136	-10.8	121	0.00
29 S	Terphenyl-d14	0.884	0.923	-4.4	121	0.00
30	Benzo(a)anthracene	1.118	1.064	4.8	110	0.00
31	Chrysene	1.196	1.170	2.2	109	0.00
32	Bis(2-ethylhexyl)phthalate	2.136	1.979	7.4	119	0.00
33	Indeno(1,2,3-cd)pyrene	1.409	1.368	2.9	108	0.00
34 I	Perylene-d12	1.000	1.000	0.0	109	0.00
35	Benzo(b)fluoranthene	1.084	0.967	10.8	108	0.00
36	Benzo(k)fluoranthene	1.232	1.135	7.9	107	0.00
37 C	Benzo(a)pyrene	0.995	0.923	7.2	110	0.00
38	Dibenzo(a,h)anthracene	1.047	0.956	8.7	108	0.00
39	Benzo(a,h,i)perylene	1.060	0.974	8.1	105	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100819\
Data File : BE100616.D
Acq On : 8 Oct 2019 19:43
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 09 01:46:28 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 09 01:42:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			