

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100819\
 Data File : BE100622.D
 Acq On : 8 Oct 2019 23:21
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 09 01:47:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 09 01:42:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	105	0.00
2	1,4-Dioxane	0.400	0.360	10.0	98	0.00
3	n-Nitrosodimethylamine	0.400	0.393	1.8	124	-0.01
4 S	2-Fluorophenol	0.400	0.412	-3.0	115	0.00
5 S	Phenol-d6	0.400	0.419	-4.7	120	0.00
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	111	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	113	0.00
8 S	Nitrobenzene-d5	0.400	0.372	7.0	117	0.00
9	Naphthalene	0.400	0.393	1.8	112	0.00
10	Hexachlorobutadiene	0.400	0.385	3.8	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.395	1.3	114	-0.01
12	2-Methylnaphthalene	0.400	0.397	0.8	116	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.400	0.388	3.0	126	0.00
15 S	2-Fluorobiphenyl	0.400	0.387	3.3	114	0.00
16	Acenaphthylene	0.400	0.372	7.0	116	0.00
17	Acenaphthene	0.400	0.386	3.5	118	0.00
18	Fluorene	0.400	0.384	4.0	119	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	118	0.00
20	4-Bromophenyl-phenylether	0.400	0.365	8.8	116	0.00
21	Hexachlorobenzene	0.400	0.380	5.0	115	0.00
22	Pentachlorophenol	0.400	0.536	-34.0#	138	-0.01
23	Phenanthrene	0.400	0.397	0.8	119	0.00
24	Anthracene	0.400	0.367	8.3	120	0.00
25 SURR	Fluoranthene-d10	0.400	0.373	6.8	120	0.00
26	Fluoranthene	0.400	0.384	4.0	120	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	108	0.01
28	Pvrene	0.400	0.449	-12.2	121	0.00
29 S	Terphenyl-d14	0.400	0.426	-6.5	121	0.00
30	Benzo(a)anthracene	0.400	0.383	4.3	109	0.00
31	Chrysene	0.400	0.399	0.3	109	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.328	18.0	103	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.390	2.5	106	0.00
34 I	Perylene-d12	0.400	0.400	0.0	108	0.00
35	Benzo(b)fluoranthene	0.400	0.343	14.2	103	0.00
36	Benzo(k)fluoranthene	0.400	0.363	9.3	104	0.00
37 C	Benzo(a)pyrene	0.400	0.365	8.8	107	0.00
38	Dibenzo(a,h)anthracene	0.400	0.362	9.5	106	0.00
39	Benzo(a,h,i)perylene	0.400	0.368	8.0	105	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE100819\
Data File : BE100622.D
Acq On : 8 Oct 2019 23:21
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 09 01:47:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 09 01:42:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				