

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100918\
 Data File : BE097814.D
 Acq On : 9 Oct 2018 12:45
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 09 14:14:09 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 09 11:59:33 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	67	-0.01
2	1,4-Dioxane	0.400	0.395	1.3	72	0.00
3	n-Nitrosodimethylamine	0.400	0.346	13.5	62	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	65	-0.01
5 S	Phenol-d6	0.400	0.341	14.8	59	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	66	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	70	0.00
8 S	Nitrobenzene-d5	0.400	0.648	-62.0#	122	0.00
9	Naphthalene	0.400	0.379	5.3	68	-0.01
10	Hexachlorobutadiene	0.400	0.401	-0.3	73	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	69	-0.01
12	2-Methylnaphthalene	0.400	0.392	2.0	72	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	74	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.396	1.0	92	-0.01
15 S	2-Fluorobiphenyl	0.400	0.369	7.8	73	-0.01
16	Acenaphthylene	0.400	0.352	12.0	69	0.00
17	Acenaphthene	0.400	0.374	6.5	73	0.00
18	Fluorene	0.400	0.371	7.3	73	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	73	-0.01
20	4-Bromophenyl-phenylether	0.400	0.387	3.3	76	-0.01
21	Hexachlorobenzene	0.400	0.377	5.8	72	0.00
22	Pentachlorophenol	0.400	0.462	-15.5	95	-0.01
23	Phenanthrene	0.400	0.377	5.8	72	0.00
24	Anthracene	0.400	0.344	14.0	69	0.00
25 SURR	Fluoranthene-d10	0.400	0.369	7.8	72	-0.01
26	Fluoranthene	0.400	0.372	7.0	73	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	67	-0.01
28	Pvrene	0.400	0.397	0.8	73	-0.01
29 S	Terphenyl-d14	0.400	0.409	-2.2	76	-0.01
30	Benzo(a)anthracene	0.400	0.362	9.5	67	-0.01
31	Chrysene	0.400	0.387	3.3	70	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.358	10.5	69	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.361	9.8	66	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	69	-0.02
35	Benzo(b)fluoranthene	0.400	0.342	14.5	66	-0.02
36	Benzo(k)fluoranthene	0.400	0.388	3.0	75	-0.02
37 C	Benzo(a)pyrene	0.400	0.332	17.0	64	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.346	13.5	67	-0.03
39	Benzo(a,h,i)perylene	0.400	0.360	10.0	68	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100918\
Data File : BE097814.D
Acq On : 9 Oct 2018 12:45
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 09 14:14:09 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 09 11:59:33 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				