

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE101118\
 Data File : BE097864.D
 Acq On : 11 Oct 2018 19:18
 Operator : SJ/JU
 Sample : J5184-12
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 JLC34

Manual Integrations
 APPROVED

Sohil
 10/12/2018 6:34:54 PM

Quant Time: Oct 11 20:03:15 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE101118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 11 14:16:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	4020	0.40	ng/ul	0.00
2) Naphthalene-d8	10.69	136	20128	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.55	164	15185m	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.31	188	31601m	0.40	ng/ul	0.01
16) Chrysene-d12	21.52	240	26688m	0.40	ng/ul	0.04
20) Perylene-d12	24.10	264	32270m	0.40	ng/ul	0.07
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.28	152	18992	0.56	ng/ul	0.00
14) Fluoranthene-d10	19.35	212	51229m	0.50	ng/ul	0.02
Target Compounds				Qvalue		
3) Naphthalene	10.74	128	150425	3.071	ng/ul	99
5) 2-Methylnaphthalene	12.36	142	67045	1.892	ng/ul	100
7) Acenaphthylene	14.27	152	44075	0.767	ng/ul#	92
8) Acenaphthene	14.61	153	156339	3.507	ng/ul	98
9) Fluorene	15.61	166	203465	3.498	ng/ul	99
11) Pentachlorophenol	16.96	266	75574m	16.964	ng/ul	
12) Phenanthrene	17.36	178	4735557m	63.299	ng/ul	
13) Anthracene	17.44	178	1623020m	23.412	ng/ul	
15) Fluoranthene	19.39	202	10991568m	114.401	ng/ul	
17) Pyrene	19.76	202	14103511	177.392	ng/ul#	90
18) Benzo(a)anthracene	21.50	228	8314561m	103.040	ng/ul	
19) Chrysene	21.56	228	11556173m	159.392	ng/ul	
21) Benzo(b)fluoranthene	23.34	252	13708942m	158.952	ng/ul	
22) Benzo(k)fluoranthene	23.38	252	3792290m	41.078	ng/ul	
23) Benzo(a)pyrene	24.01	252	6459544m	76.286	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.86	276	4856706	50.525	ng/ul#	100
25) Dibenzo(a,h)anthracene	26.88	278	1372331	17.274	ng/ul	96
26) Benzo(g,h,i)perylene	27.71	276	4305761	51.409	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE101118\
Data File : BE097864.D
Acq On : 11 Oct 2018 19:18
Operator : SJ/JU
Sample : J5184-12
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampled :
JLC34

Manual Integrations
APPROVED

Sohil
10/12/2018 6:34:54 PM

Quant Time: Oct 11 20:03:15 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE101118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 11 14:16:20 2018
Response via : Initial Calibration

