

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE101118\
 Data File : BE097876.D
 Acq On : 12 Oct 2018 2:43
 Operator : SJ/JU
 Sample : J5184-12DL 5X
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 JLC34DL

Manual Integrations
 APPROVED

Sohil

10/12/2018 6:35:05 PM

Quant Time: Oct 12 05:22:12 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE101118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 12 04:56:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	2922	0.40	ng/ul	0.00
2) Naphthalene-d8	10.70	136	15025	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.56	164	10432	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.31	188	24158	0.40	ng/ul	0.00
16) Chrysene-d12	21.50	240	23343	0.40	ng/ul	0.00
20) Perylene-d12	24.07	264	27641	0.40	ng/ul	0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.29	152	2875	0.11	ng/ul	0.00
14) Fluoranthene-d10	19.34	212	8070	0.10	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.75	128	22483	0.615	ng/ul	98
5) 2-Methylnaphthalene	12.36	142	10144	0.384	ng/ul	98
7) Acenaphthylene	14.27	152	6720	0.170	ng/ul#	88
8) Acenaphthene	14.61	153	24123	0.788	ng/ul	97
9) Fluorene	15.61	166	32018	0.801	ng/ul	98
11) Pentachlorophenol	16.97	266	8600	2.525	ng/ul	98
12) Phenanthrene	17.35	178	928994	16.243	ng/ul	98
13) Anthracene	17.44	178	265823	5.016	ng/ul	99
15) Fluoranthene	19.38	202	2627857	35.778	ng/ul	95
17) Pyrene	19.74	202	2816087	40.496	ng/ul	96
18) Benzo(a)anthracene	21.49	228	2087460	29.576	ng/ul	97
19) Chrysene	21.54	228	2763574	43.580	ng/ul	93
21) Benzo(b)fluoranthene	23.29	252	2974393m	40.263	ng/ul	
22) Benzo(k)fluoranthene	23.34	252	777472m	9.832	ng/ul	
23) Benzo(a)pyrene	23.96	252	1373887	18.942	ng/ul	99
24) Indeno(1,2,3-cd)pyrene	26.79	276	939294	11.408	ng/ul#	100
25) Dibenzo(a,h)anthracene	26.81	278	275186	4.044	ng/ul#	94
26) Benzo(g,h,i)perylene	27.63	276	841483	11.729	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE101118\
Data File : BE097876.D
Acq On : 12 Oct 2018 2:43
Operator : SJ/JU
Sample : J5184-12DL 5X
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
JLC34DL

Manual Integrations
APPROVED

Sohil

10/12/2018 6:35:05 PM

Quant Time: Oct 12 05:22:12 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE101118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 12 04:56:57 2018
Response via : Initial Calibration

