

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE101118\
 Data File : BE097857.D
 Acq On : 11 Oct 2018 14:59
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.451

Quant Time: Oct 11 16:30:34 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE101118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 11 14:16:20 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	102	-0.01
3	Naphthalene	0.973	0.963	1.0	102	0.00
4 SURR	2-Methylnaphthalene-d10	0.679	0.669	1.5	102	0.00
5	2-Methylnaphthalene	0.704	0.683	3.0	101	-0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	100	-0.01
7	Acenaphthylene	1.513	1.509	0.3	100	-0.01
8 C	Acenaphthene	1.174	1.181	-0.6	100	-0.01
9	Fluorene	1.532	1.526	0.4	100	-0.02
10 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.01
11	Pentachlorophenol	0.056	0.048	14.3	102	-0.03
12	Phenanthrene	0.947	0.947	0.0	99	-0.01
13	Anthracene	0.877	0.854	2.6	98	-0.01
14 SURR	Fluoranthene-d10	1.292	1.262	2.3	98	-0.01
15 C	Fluoranthene	1.216	1.203	1.1	99	-0.01
16 I	Chrysene-d12	1.000	1.000	0.0	99	-0.01
17	Pyrene	1.192	1.238	-3.9	95	-0.01
18	Benzo(a)anthracene	1.209	1.287	-6.5	103	-0.01
19	Chrysene	1.087	1.144	-5.2	98	-0.02
20 I	Perylene-d12	1.000	1.000	0.0	98	-0.03
21	Benzo(b)fluoranthene	1.069	1.034	3.3	97	-0.02
22	Benzo(k)fluoranthene	1.144	1.127	1.5	98	-0.03
23 C	Benzo(a)pyrene	1.050	1.003	4.5	96	-0.03
24	Indeno(1,2,3-cd)pyrene	1.192	1.148	3.7	97	-0.05
25	Dibenzo(a,h)anthracene	0.985	0.960	2.5	98	-0.04
26	Benzo(g,h,i)perylene	1.038	1.011	2.6	98	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0