

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101119\
 Data File : BE100633.D
 Acq On : 11 Oct 2019 9:27
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 11 14:30:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 10 14:01:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.714	0.688	3.6	97	0.00
3	n-Nitrosodimethylamine	0.755	0.656	13.1	95	0.00
4 S	2-Fluorophenol	1.235	1.132	8.3	97	0.00
5 S	Phenol-d6	1.742	1.564	10.2	95	0.00
6	bis(2-Chloroethyl)ether	1.430	1.394	2.5	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.248	0.231	6.9	105	0.00
9	Naphthalene	4.206	4.200	0.1	101	0.00
10	Hexachlorobutadiene	0.183	0.179	2.2	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.631	0.610	3.3	100	0.01
12	2-Methylnaphthalene	0.717	0.690	3.8	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.106	0.087	17.9	91	0.00
15 S	2-Fluorobiphenyl	1.457	1.465	-0.5	100	0.00
16	Acenaphthylene	1.721	1.549	10.0	91	0.00
17	Acenaphthene	1.112	1.078	3.1	98	0.00
18	Fluorene	1.477	1.421	3.8	97	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4-Bromophenyl-phenylether	0.204	0.193	5.4	99	0.00
21	Hexachlorobenzene	0.194	0.186	4.1	102	0.00
22	Pentachlorophenol	0.062	0.064	-3.2	116	0.00
23	Phenanthrene	1.100	1.132	-2.9	102	0.00
24	Anthracene	1.021	0.935	8.4	95	0.00
25 SURR	Fluoranthene-d10	5.162	5.099	1.2	103	0.00
26	Fluoranthene	1.436	1.481	-3.1	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
28	Pyrene	1.118	1.055	5.6	103	0.00
29 S	Terphenyl-d14	0.875	0.814	7.0	107	0.00
30	Benzo(a)anthracene	1.305	1.221	6.4	103	0.00
31	Chrysene	1.304	1.301	0.2	109	0.00
32	Bis(2-ethylhexyl)phthalate	2.729	1.917	29.8#	76	0.00
33	Indeno(1,2,3-cd)pyrene	1.603	1.386	13.5	97	0.00
34 I	Perylene-d12	1.000	1.000	0.0	101	0.00
35	Benzo(b)fluoranthene	1.166	1.057	9.3	99	0.00
36	Benzo(k)fluoranthene	1.200	1.095	8.8	98	0.00
37 C	Benzo(a)pyrene	1.086	0.950	12.5	96	0.00
38	Dibenzo(a,h)anthracene	1.122	0.978	12.8	97	0.00
39	Benzo(a,h,i)perylene	1.129	1.017	9.9	98	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101119\
Data File : BE100633.D
Acq On : 11 Oct 2019 9:27
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 11 14:30:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 10 14:01:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0		CCC's out = 0	