

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101119\
 Data File : BE100639.D
 Acq On : 11 Oct 2019 17:21
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 11 18:35:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 10 14:01:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	-0.01
2	1,4-Dioxane	0.714	0.712	0.3	80	0.00
3	n-Nitrosodimethylamine	0.755	0.771	-2.1	88	-0.01
4 S	2-Fluorophenol	1.235	1.516	-22.8	102	0.00
5 S	Phenol-d6	1.742	2.207	-26.7#	106	0.00
6	bis(2-Chloroethyl)ether	1.430	1.387	3.0	82	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	80	0.00
8 S	Nitrobenzene-d5	0.248	0.339	-36.7#	124	0.00
9	Naphthalene	4.206	4.150	1.3	81	0.00
10	Hexachlorobutadiene	0.183	0.189	-3.3	86	0.00
11 SURR	2-Methylnaphthalene-d10	0.631	0.600	4.9	79	0.00
12	2-Methylnaphthalene	0.717	0.699	2.5	81	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.00
14 S	2,4,6-Tribromophenol	0.106	0.143	-34.9#	121	0.00
15 S	2-Fluorobiphenyl	1.457	1.885	-29.4#	104	0.00
16	Acenaphthylene	1.721	1.551	9.9	74	0.00
17	Acenaphthene	1.112	1.123	-1.0	83	0.00
18	Fluorene	1.477	1.409	4.6	78	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	81	0.00
20	4-Bromophenyl-phenylether	0.204	0.191	6.4	79	-0.01
21	Hexachlorobenzene	0.194	0.193	0.5	85	0.00
22	Pentachlorophenol	0.062	0.078	-25.8#	113	0.00
23	Phenanthrene	1.100	1.120	-1.8	82	0.00
24	Anthracene	1.021	0.935	8.4	77	0.00
25 SURR	Fluoranthene-d10	5.162	5.046	2.2	82	0.00
26	Fluoranthene	1.436	1.503	-4.7	85	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
28	Pvrene	1.118	1.033	7.6	85	0.00
29 S	Terphenyl-d14	0.875	1.045	-19.4	115	0.00
30	Benzo(a)anthracene	1.305	1.230	5.7	87	0.00
31	Chrysene	1.304	1.295	0.7	91	0.00
32	Bis(2-ethylhexyl)phthalate	2.729	1.895	30.6#	63	0.00
33	Indeno(1,2,3-cd)pyrene	1.603	1.388	13.4	82	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	85	0.00
35	Benzo(b)fluoranthene	1.166	1.068	8.4	84	0.00
36	Benzo(k)fluoranthene	1.200	1.125	6.2	85	0.00
37 C	Benzo(a)pyrene	1.086	0.933	14.1	79	0.00
38	Dibenzo(a,h)anthracene	1.122	0.969	13.6	81	-0.01
39	Benzo(a,h,i)perylene	1.129	1.032	8.6	83	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101119\
Data File : BE100639.D
Acq On : 11 Oct 2019 17:21
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 11 18:35:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 10 14:01:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			