

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE101218\
 Data File : BE097888.D
 Acq On : 12 Oct 2018 20:40
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.454

Quant Time: Oct 13 02:58:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE101118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 13 02:36:51 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
3	Naphthalene	0.973	1.061	-9.0	103	0.00
4 SURR	2-Methylnaphthalene-d10	0.679	0.736	-8.4	102	0.00
5	2-Methylnaphthalene	0.704	0.765	-8.7	103	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.00
7	Acenaphthylene	1.513	1.786	-18.0	95	0.00
8 C	Acenaphthene	1.174	1.299	-10.6	89	0.00
9	Fluorene	1.532	1.533	-0.1	80	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	60	0.00
11	Pentachlorophenol	0.056	0.038	32.1#	49	0.00
12	Phenanthrene	0.947	1.048	-10.7	67	0.00
13	Anthracene	0.877	1.035	-18.0	72	0.00
14 SURR	Fluoranthene-d10	1.292	1.387	-7.4	66	0.00
15 C	Fluoranthene	1.216	1.350	-11.0	68	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	56	0.00
17	Pyrene	1.192	1.505	-26.3#	65	0.00
18	Benzo(a)anthracene	1.209	1.348	-11.5	61	0.00
19	Chrysene	1.087	1.355	-24.7#	66	0.00
20 I	Perylene-d12	1.000	1.000	0.0	55	0.00
21	Benzo(b)fluoranthene	1.069	1.298	-21.4#	69	0.00
22	Benzo(k)fluoranthene	1.144	1.235	-8.0	60	0.00
23 C	Benzo(a)pyrene	1.050	1.253	-19.3	67	0.00
24	Indeno(1,2,3-cd)pyrene	1.192	1.397	-17.2	66	0.00
25	Dibenzo(a,h)anthracene	0.985	1.182	-20.0	67	0.00
26	Benzo(g,h,i)perylene	1.038	1.135	-9.3	61	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0