

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE101218\
 Data File : BE097898.D
 Acq On : 13 Oct 2018 2:50
 Operator : SJ/JU
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.455

Quant Time: Oct 13 04:29:57 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE101118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 13 02:57:21 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
3	Naphthalene	0.973	1.052	-8.1	109	0.00
4 SURR	2-Methylnaphthalene-d10	0.679	0.727	-7.1	108	0.00
5	2-Methylnaphthalene	0.704	0.764	-8.5	110	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
7	Acenaphthylene	1.513	1.776	-17.4	99	0.00
8 C	Acenaphthene	1.174	1.307	-11.3	94	0.00
9	Fluorene	1.532	1.534	-0.1	85	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	63	0.00
11	Pentachlorophenol	0.056	0.027	51.8#	36	0.00
12	Phenanthrene	0.947	1.044	-10.2	70	0.00
13	Anthracene	0.877	1.054	-20.2#	77	0.00
14 SURR	Fluoranthene-d10	1.292	1.397	-8.1	69	0.00
15 C	Fluoranthene	1.216	1.338	-10.0	70	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	56	0.00
17	Pyrene	1.192	1.654	-38.8#	73	0.00
18	Benzo(a)anthracene	1.209	1.344	-11.2	61	0.00
19	Chrysene	1.087	1.348	-24.0#	66	0.00
20 I	Perylene-d12	1.000	1.000	0.0	56	0.00
21	Benzo(b)fluoranthene	1.069	1.307	-22.3#	71	0.00
22	Benzo(k)fluoranthene	1.144	1.242	-8.6	62	0.00
23 C	Benzo(a)pyrene	1.050	1.242	-18.3	68	0.00
24	Indeno(1,2,3-cd)pyrene	1.192	1.385	-16.2	67	0.00
25	Dibenzo(a,h)anthracene	0.985	1.174	-19.2	69	0.00
26	Benzo(g,h,i)perylene	1.038	1.104	-6.4	61	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0