

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_E\DATA\BE101218\  
 Data File : BE097887.D  
 Acq On : 12 Oct 2018 19:25  
 Operator : SJ/JU  
 Sample : J5183-07  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 JLC45

Manual Integrations  
 APPROVED

Sohil  
 10/15/2018 3:55:43 PM

Quant Time: Oct 13 02:46:00 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\SOM-EPA-SIM-BE101118.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Sat Oct 13 02:36:51 2018  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.89  | 152  | 1935     | 0.40 | ng/ul | 0.00     |
| 2) Naphthalene-d8         | 10.70 | 136  | 10047    | 0.40 | ng/ul | 0.00     |
| 6) Acenaphthene-d10       | 14.56 | 164  | 7006     | 0.40 | ng/ul | 0.00     |
| 10) Phenanthrene-d10      | 17.32 | 188  | 15247m   | 0.40 | ng/ul | 0.00     |
| 16) Chrysene-d12          | 21.50 | 240  | 16437m   | 0.40 | ng/ul | 0.00     |
| 20) Perylene-d12          | 24.08 | 264  | 15239m   | 0.40 | ng/ul | 0.02     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|-------|------|----------|------|-------|----------|
| 4) 2-Methylnaphthalene-d10  | 12.30 | 152  | 6564     | 0.38 | ng/ul | 0.00     |
| 14) Fluoranthene-d10        | 19.35 | 212  | 19467m   | 0.40 | ng/ul | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_E\DATA\BE101218\  
Data File : BE097887.D  
Acq On : 12 Oct 2018 19:25  
Operator : SJ/JU  
Sample : J5183-07  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
JLC45

Manual Integrations  
APPROVED

Sohil  
10/15/2018 3:55:43 PM

Quant Time: Oct 13 02:46:00 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\SOM-EPA-SIM-BE101118.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat Oct 13 02:36:51 2018  
Response via : Initial Calibration

