

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE101518\
 Data File : BE097903.D
 Acq On : 15 Oct 2018 17:18
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.457

Quant Time: Oct 15 17:48:18 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE101118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 15 16:15:33 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	92	0.01
3	Naphthalene	0.973	1.055	-8.4	101	0.01
4 SURR	2-Methylnaphthalene-d10	0.679	0.730	-7.5	100	0.01
5	2-Methylnaphthalene	0.704	0.757	-7.5	101	0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.01
7	Acenaphthylene	1.513	1.805	-19.3	96	0.01
8 C	Acenaphthene	1.174	1.300	-10.7	89	0.00
9	Fluorene	1.532	1.513	1.2	79	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	61	0.00
11	Pentachlorophenol	0.056	0.076	-35.7#	99	0.00
12	Phenanthrene	0.947	1.036	-9.4	67	0.00
13	Anthracene	0.877	1.044	-19.0	74	0.00
14 SURR	Fluoranthene-d10	1.292	1.377	-6.6	66	0.00
15 C	Fluoranthene	1.216	1.294	-6.4	66	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	52	0.00
17	Pyrene	1.192	1.577	-32.3#	64	0.00
18	Benzo(a)anthracene	1.209	1.340	-10.8	56	0.00
19	Chrysene	1.087	1.302	-19.8	58	0.00
20 I	Perylene-d12	1.000	1.000	0.0	52	0.00
21	Benzo(b)fluoranthene	1.069	1.252	-17.1	63	0.00
22	Benzo(k)fluoranthene	1.144	1.268	-10.8	59	0.00
23 C	Benzo(a)pyrene	1.050	1.220	-16.2	63	0.00
24	Indeno(1,2,3-cd)pyrene	1.192	1.367	-14.7	62	0.00
25	Dibenzo(a,h)anthracene	0.985	1.148	-16.5	63	0.00
26	Benzo(g,h,i)perylene	1.038	1.076	-3.7	56	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0