

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101519\
 Data File : BE100641.D
 Acq On : 15 Oct 2019 9:06
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 15 09:50:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 10 14:01:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	-0.01
2	1,4-Dioxane	0.714	0.669	6.3	85	0.00
3	n-Nitrosodimethylamine	0.755	0.708	6.2	91	-0.01
4 S	2-Fluorophenol	1.235	1.189	3.7	91	0.00
5 S	Phenol-d6	1.742	1.737	0.3	95	0.00
6	bis(2-Chloroethyl)ether	1.430	1.353	5.4	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.248	0.268	-8.1	114	0.00
9	Naphthalene	4.206	4.126	1.9	93	-0.01
10	Hexachlorobutadiene	0.183	0.183	0.0	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.631	0.602	4.6	92	0.00
12	2-Methylnaphthalene	0.717	0.678	5.4	91	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	88	-0.01
14 S	2,4,6-Tribromophenol	0.106	0.097	8.5	91	0.00
15 S	2-Fluorobiphenyl	1.457	1.637	-12.4	101	0.00
16	Acenaphthylene	1.721	1.502	12.7	80	0.00
17	Acenaphthene	1.112	1.097	1.3	90	0.00
18	Fluorene	1.477	1.410	4.5	87	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
20	4-Bromophenyl-phenylether	0.204	0.191	6.4	89	-0.01
21	Hexachlorobenzene	0.194	0.191	1.5	95	0.00
22	Pentachlorophenol	0.062	0.068	-9.7	112	0.00
23	Phenanthrene	1.100	1.120	-1.8	93	0.00
24	Anthracene	1.021	0.904	11.5	84	0.00
25 SURR	Fluoranthene-d10	5.162	5.127	0.7	95	0.00
26	Fluoranthene	1.436	1.510	-5.2	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
28	Pvrene	1.118	1.069	4.4	97	0.00
29 S	Terphenyl-d14	0.875	0.949	-8.5	115	0.00
30	Benzo(a)anthracene	1.305	1.174	10.0	92	0.00
31	Chrysene	1.304	1.268	2.8	99	0.00
32	Bis(2-ethylhexyl)phthalate	2.729	1.373	49.7#	51	-0.01
33	Indeno(1,2,3-cd)pyrene	1.603	1.352	15.7	88	0.00
34 I	Perylene-d12	1.000	1.000	0.0	93	0.00
35	Benzo(b)fluoranthene	1.166	1.051	9.9	90	0.00
36	Benzo(k)fluoranthene	1.200	1.103	8.1	91	0.00
37 C	Benzo(a)pyrene	1.086	0.910	16.2	84	0.00
38	Dibenzo(a,h)anthracene	1.122	0.977	12.9	90	-0.01
39	Benzo(a,h,i)perylene	1.129	1.025	9.2	91	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101519\
Data File : BE100641.D
Acq On : 15 Oct 2019 9:06
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 15 09:50:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 10 14:01:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			