

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101519\
 Data File : BE100651.D
 Acq On : 15 Oct 2019 16:46
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 15 18:49:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 10 14:01:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2	1,4-Dioxane	0.714	0.554	22.4	63	0.00
3	n-Nitrosodimethylamine	0.755	0.668	11.5	77	0.00
4 S	2-Fluorophenol	1.235	1.302	-5.4	89	0.00
5 S	Phenol-d6	1.742	2.091	-20.0	101	0.00
6	bis(2-Chloroethyl)ether	1.430	1.384	3.2	82	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
8 S	Nitrobenzene-d5	0.248	0.365	-47.2#	150#	0.00
9	Naphthalene	4.206	4.031	4.2	88	0.00
10	Hexachlorobutadiene	0.183	0.177	3.3	90	0.00
11 SURR	2-Methylnaphthalene-d10	0.631	0.606	4.0	90	0.01
12	2-Methylnaphthalene	0.717	0.701	2.2	91	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
14 S	2,4,6-Tribromophenol	0.106	0.159	-50.0#	150#	0.00
15 S	2-Fluorobiphenyl	1.457	1.566	-7.5	96	0.00
16	Acenaphthylene	1.721	1.749	-1.6	93	0.00
17	Acenaphthene	1.112	1.079	3.0	88	0.00
18	Fluorene	1.477	1.408	4.7	87	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	81	0.00
20	4-Bromophenyl-phenylether	0.204	0.209	-2.5	86	0.00
21	Hexachlorobenzene	0.194	0.193	0.5	85	0.01
22	Pentachlorophenol	0.062	0.102	-64.5#	149	0.01
23	Phenanthrene	1.100	1.154	-4.9	84	0.00
24	Anthracene	1.021	1.069	-4.7	88	0.00
25 SURR	Fluoranthene-d10	5.162	4.665	9.6	76	0.00
26	Fluoranthene	1.436	1.435	0.1	81	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	63	0.01
28	Pvrene	1.118	1.446	-29.3#	82	0.00
29 S	Terphenyl-d14	0.875	1.095	-25.1#	83	0.00
30	Benzo(a)anthracene	1.305	1.393	-6.7	68	0.00
31	Chrysene	1.304	1.360	-4.3	66	0.01
32	Bis(2-ethylhexyl)phthalate	2.729	3.478	-27.4#	80	0.00
33	Indeno(1,2,3-cd)pyrene	1.603	1.486	7.3	60	0.01
34 I	Perylene-d12	1.000	1.000	0.0	57	0.00
35	Benzo(b)fluoranthene	1.166	1.189	-2.0	63	0.00
36	Benzo(k)fluoranthene	1.200	1.161	3.2	59	0.00
37 C	Benzo(a)pyrene	1.086	1.126	-3.7	65	0.00
38	Dibenzo(a,h)anthracene	1.122	1.069	4.7	61	0.01
39	Benzo(a,h,i)perylene	1.129	1.022	9.5	56	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101519\
Data File : BE100651.D
Acq On : 15 Oct 2019 16:46
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 15 18:49:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 10 14:01:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			