

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_E\DATA\BE101618\  
 Data File : BE097920.D  
 Acq On : 16 Oct 2018 21:18  
 Operator : SJ/JU  
 Sample : J5393-07  
 Misc :  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 RE-123D3-20181008

Quant Time: Oct 17 07:43:50 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE101618.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Oct 16 15:36:44 2018  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|--------------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4      | 7.88  | 152  | 1876     | 0.40   | ng    | 0.00     |
| 7) Naphthalene-d8              | 10.69 | 136  | 8263     | 0.40   | ng    | 0.00     |
| 13) Acenaphthene-d10           | 14.56 | 164  | 1568     | 0.40   | ng    | 0.00     |
| 19) Phenanthrene-d10           | 17.29 | 188  | 10133    | 0.40   | ng    | -0.01    |
| 27) Chrysene-d12               | 21.49 | 240  | 6159     | 0.40   | ng    | 0.00     |
| 34) Perylene-d12               | 24.05 | 264  | 5        | 0.40   | ng    | 0.00     |
| System Monitoring Compounds    |       |      |          |        |       |          |
| 4) 2-Fluorophenol              | 5.47  | 112  | 711      | 0.12   | ng    | 0.00     |
| 5) Phenol-d6                   | 7.06  | 99   | 350      | 0.04   | ng    | 0.00     |
| 8) Nitrobenzene-d5             | 9.03  | 82   | 2619     | 0.33   | ng    | -0.01    |
| 11) 2-Methylnaphthalene-d10    | 12.28 | 152  | 3758     | 0.27   | ng    | -0.01    |
| 14) 2,4,6-Tribromophenol       | 16.05 | 330  | 878      | 1.04   | ng    | 0.00     |
| 15) 2-Fluorobiphenyl           | 13.17 | 172  | 7657     | 1.19   | ng    | 0.00     |
| 25) Fluoranthene-d10           | 19.33 | 212  | 22855    | 0.17   | ng    | 0.00     |
| 29) Terphenyl-d14              | 19.92 | 244  | 15711    | 1.11   | ng    | 0.00     |
| Target Compounds               |       |      |          |        |       |          |
| 2) 1,4-Dioxane                 | 3.38  | 88   | 878      | 0.268  | ng    | # 87     |
| 32) Bis(2-ethylhexyl)phthalate | 21.39 | 149  | 16604    | 0.369  | ng    | # 99     |
| 35) Benzo(b)fluoranthene       | 23.29 | 252  | 343      | 23.839 | ng    | # 1      |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_E\DATA\BE101618\  
Data File : BE097920.D  
Acq On : 16 Oct 2018 21:18  
Operator : SJ/JU  
Sample : J5393-07  
Misc :  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
RE-123D3-20181008

Quant Time: Oct 17 07:43:50 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE101618.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Oct 16 15:36:44 2018  
Response via : Initial Calibration

