

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE101618\
 Data File : BE097944.D
 Acq On : 17 Oct 2018 12:43
 Operator : SJ/JU
 Sample : J5317-06DL 2X
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE-122D2-20181004DL

Quant Time: Oct 17 17:17:53 2018

Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101618.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Oct 16 15:36:44 2018

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	1992	0.40	ng	0.00
7) Naphthalene-d8	10.68	136	8234	0.40	ng	-0.01
13) Acenaphthene-d10	14.54	164	5163	0.40	ng	-0.01
19) Phenanthrene-d10	17.29	188	14217	0.40	ng	-0.01
27) Chrysene-d12	21.47	240	17166	0.40	ng	-0.01
34) Perylene-d12	24.03	264	17667	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.47	112	565	0.09	ng	0.00
5) Phenol-d6	7.06	99	495	0.05	ng	0.00
8) Nitrobenzene-d5	9.04	82	1604	0.21	ng	0.00
11) 2-Methylnaphthalene-d10	12.28	152	2526	0.18	ng	-0.01
14) 2,4,6-Tribromophenol	16.05	330	418	0.15	ng	0.00
15) 2-Fluorobiphenyl	13.17	172	4957	0.23	ng	0.00
25) Fluoranthene-d10	19.33	212	36322	0.19	ng	0.00
29) Terphenyl-d14	19.92	244	9889	0.25	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.38	88	8842	2.546	ng	93
6) bis(2-Chloroethyl)ether	7.30	93	220	0.028	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE101618\
Data File : BE097944.D
Acq On : 17 Oct 2018 12:43
Operator : SJ/JU
Sample : J5317-06DL 2X
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE-122D2-20181004DL

Quant Time: Oct 17 17:17:53 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

