

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101618\
 Data File : BE097914.D
 Acq On : 16 Oct 2018 17:38
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 17 02:19:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.697	0.781	-12.1	104	0.00
3	n-Nitrosodimethylamine	0.779	0.700	10.1	87	0.00
4 S	2-Fluorophenol	1.261	1.264	-0.2	91	0.00
5 S	Phenol-d6	1.864	1.800	3.4	85	0.00
6	bis(2-Chloroethyl)ether	1.583	1.488	6.0	86	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	84	0.00
8 S	Nitrobenzene-d5	0.379	0.363	4.2	82	0.00
9	Naphthalene	3.983	4.117	-3.4	89	0.00
10	Hexachlorobutadiene	0.215	0.226	-5.1	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.665	0.0	86	0.00
12	2-Methylnaphthalene	0.682	0.670	1.8	86	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	82	0.00
14 S	2,4,6-Tribromophenol	0.215	0.235	-9.3	102	0.00
15 S	2-Fluorobiphenyl	1.635	1.593	2.6	85	0.00
16	Acenaphthylene	1.848	1.816	1.7	85	0.00
17	Acenaphthene	1.129	1.125	0.4	86	0.00
18	Fluorene	1.554	1.562	-0.5	87	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
20	4-Bromophenyl-phenylether	0.219	0.227	-3.7	89	0.00
21	Hexachlorobenzene	0.215	0.222	-3.3	88	0.01
22	Pentachlorophenol	0.062	0.049	21.0	87	0.00
23	Phenanthrene	1.016	1.044	-2.8	89	0.00
24	Anthracene	0.960	0.965	-0.5	88	0.00
25 SURR	Fluoranthene-d10	5.421	5.672	-4.6	94	0.00
26	Fluoranthene	1.334	1.413	-5.9	95	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
28	Pvrene	1.154	1.106	4.2	97	0.00
29 S	Terphenyl-d14	0.917	1.179	-28.6#	135	0.00
30	Benzo(a)anthracene	1.227	1.253	-2.1	104	0.00
31	Chrysene	1.162	1.195	-2.8	107	0.00
32	Bis(2-ethylhexyl)phthalate	2.865	2.890	-0.9	99	0.00
33	Indeno(1,2,3-cd)pyrene	1.249	1.294	-3.6	100	0.00
34 I	Perylene-d12	1.000	1.000	0.0	94	0.00
35	Benzo(b)fluoranthene	1.151	1.155	-0.3	102	0.00
36	Benzo(k)fluoranthene	1.199	1.209	-0.8	105	0.00
37 C	Benzo(a)pyrene	1.122	1.159	-3.3	103	0.00
38	Dibenzo(a,h)anthracene	1.099	1.086	1.2	99	0.00
39	Benzo(a,h,i)perylene	1.126	1.103	2.0	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101618\
Data File : BE097914.D
Acq On : 16 Oct 2018 17:38
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 17 02:19:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			