

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101618\
 Data File : BE097932.D
 Acq On : 17 Oct 2018 5:19
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 17 08:28:04 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	134	0.00
2	1,4-Dioxane	0.697	0.667	4.3	139	0.00
3	n-Nitrosodimethylamine	0.779	0.679	12.8	132	0.00
4 S	2-Fluorophenol	1.261	1.221	3.2	137	0.00
5 S	Phenol-d6	1.864	1.855	0.5	137	0.00
6	bis(2-Chloroethyl)ether	1.583	1.578	0.3	142	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	143	-0.01
8 S	Nitrobenzene-d5	0.379	0.366	3.4	141	-0.01
9	Naphthalene	3.983	4.003	-0.5	148	-0.01
10	Hexachlorobutadiene	0.215	0.211	1.9	145	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.682	-2.6	150#	-0.02
12	2-Methylnaphthalene	0.682	0.706	-3.5	154#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	156#	0.00
14 S	2,4,6-Tribromophenol	0.215	0.213	0.9	174#	0.00
15 S	2-Fluorobiphenyl	1.635	1.526	6.7	154#	0.00
16	Acenaphthylene	1.848	1.829	1.0	163#	0.00
17	Acenaphthene	1.129	1.114	1.3	162#	0.00
18	Fluorene	1.554	1.604	-3.2	169#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	180#	-0.01
20	4-Bromophenyl-phenylether	0.219	0.206	5.9	174#	0.00
21	Hexachlorobenzene	0.215	0.205	4.7	176#	0.00
22	Pentachlorophenol	0.062	0.074	-19.4	284#	-0.01
23	Phenanthrene	1.016	0.994	2.2	182#	0.00
24	Anthracene	0.960	0.932	2.9	183#	0.00
25 SURR	Fluoranthene-d10	5.421	5.516	-1.8	197#	0.00
26	Fluoranthene	1.334	1.390	-4.2	201#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	187#	-0.01
28	Pvrene	1.154	1.174	-1.7	203#	0.00
29 S	Terphenyl-d14	0.917	1.527	-66.5#	344#	-0.01
30	Benzo(a)anthracene	1.227	1.187	3.3	192#	0.00
31	Chrysene	1.162	1.178	-1.4	208#	0.00
32	Bis(2-ethylhexyl)phthalate	2.865	2.700	5.8	181#	0.00
33	Indeno(1,2,3-cd)pyrene	1.249	0.912	27.0#	138	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	145	-0.01
35	Benzo(b)fluoranthene	1.151	1.175	-2.1	159#	0.00
36	Benzo(k)fluoranthene	1.199	1.252	-4.4	167#	0.00
37 C	Benzo(a)pyrene	1.122	1.112	0.9	152#	-0.01
38	Dibenzo(a,h)anthracene	1.099	0.999	9.1	140	-0.01
39	Benzo(a,h,i)perylene	1.126	0.985	12.5	139	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101618\
Data File : BE097932.D
Acq On : 17 Oct 2018 5:19
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 17 08:28:04 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			