

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101618\
 Data File : BE097932.D
 Acq On : 17 Oct 2018 5:19
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 17 08:28:04 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	134	0.00
2	1,4-Dioxane	0.400	0.383	4.3	139	0.00
3	n-Nitrosodimethylamine	0.400	0.349	12.8	132	0.00
4 S	2-Fluorophenol	0.400	0.387	3.3	137	0.00
5 S	Phenol-d6	0.400	0.398	0.5	137	0.00
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	142	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	143	-0.01
8 S	Nitrobenzene-d5	0.400	0.386	3.5	141	-0.01
9	Naphthalene	0.400	0.402	-0.5	148	-0.01
10	Hexachlorobutadiene	0.400	0.393	1.8	145	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	150	-0.02
12	2-Methylnaphthalene	0.400	0.414	-3.5	154	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	156	0.00
14 S	2,4,6-Tribromophenol	0.400	0.395	1.3	174	0.00
15 S	2-Fluorobiphenyl	0.400	0.373	6.8	154	0.00
16	Acenaphthylene	0.400	0.396	1.0	163	0.00
17	Acenaphthene	0.400	0.395	1.3	162	0.00
18	Fluorene	0.400	0.413	-3.2	169	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	180	-0.01
20	4-Bromophenyl-phenylether	0.400	0.377	5.8	174	0.00
21	Hexachlorobenzene	0.400	0.381	4.8	176	0.00
22	Pentachlorophenol	0.400	0.540	-35.0#	284	-0.01
23	Phenanthrene	0.400	0.391	2.3	182	0.00
24	Anthracene	0.400	0.388	3.0	183	0.00
25 SURR	Fluoranthene-d10	0.400	0.407	-1.7	197	0.00
26	Fluoranthene	0.400	0.417	-4.2	201	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	187	-0.01
28	Pvrene	0.400	0.407	-1.7	203	0.00
29 S	Terphenyl-d14	0.400	0.666	-66.5#	344	-0.01
30	Benzo(a)anthracene	0.400	0.387	3.3	192	0.00
31	Chrysene	0.400	0.406	-1.5	208	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.370	7.5	181	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.237	40.8#	138	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	145	-0.01
35	Benzo(b)fluoranthene	0.400	0.408	-2.0	159	0.00
36	Benzo(k)fluoranthene	0.400	0.418	-4.5	167	0.00
37 C	Benzo(a)pyrene	0.400	0.397	0.8	152	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.364	9.0	140	-0.01
39	Benzo(a,h,i)perylene	0.400	0.350	12.5	139	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101618\
Data File : BE097932.D
Acq On : 17 Oct 2018 5:19
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 17 08:28:04 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				