

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101618\
 Data File : BE097948.D
 Acq On : 17 Oct 2018 15:49
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 17 16:20:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.697	0.697	0.0	94	0.00
3	n-Nitrosodimethylamine	0.779	0.618	20.7	78	0.00
4 S	2-Fluorophenol	1.261	1.190	5.6	86	-0.01
5 S	Phenol-d6	1.864	1.631	12.5	78	-0.01
6	bis(2-Chloroethyl)ether	1.583	1.381	12.8	81	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	-0.01
8 S	Nitrobenzene-d5	0.379	0.333	12.1	77	-0.01
9	Naphthalene	3.983	3.749	5.9	83	-0.01
10	Hexachlorobutadiene	0.215	0.200	7.0	83	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.616	7.4	82	-0.01
12	2-Methylnaphthalene	0.682	0.606	11.1	79	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	-0.01
14 S	2,4,6-Tribromophenol	0.215	0.164	23.7	75	-0.01
15 S	2-Fluorobiphenyl	1.635	1.424	12.9	80	0.00
16	Acenaphthylene	1.848	1.660	10.2	82	0.00
17	Acenaphthene	1.129	1.031	8.7	83	0.00
18	Fluorene	1.554	1.421	8.6	83	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	-0.01
20	4-Bromophenyl-phenylether	0.219	0.189	13.7	84	0.00
21	Hexachlorobenzene	0.215	0.190	11.6	86	0.00
22	Pentachlorophenol	0.062	0.041	33.9#	82	-0.01
23	Phenanthrene	1.016	0.907	10.7	88	0.00
24	Anthracene	0.960	0.818	14.8	85	-0.01
25 SURR	Fluoranthene-d10	5.421	4.764	12.1	90	0.00
26	Fluoranthene	1.334	1.182	11.4	90	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	-0.01
28	Pvrene	1.154	1.005	12.9	95	0.00
29 S	Terphenyl-d14	0.917	0.766	16.5	94	0.00
30	Benzo(a)anthracene	1.227	1.076	12.3	96	0.00
31	Chrysene	1.162	1.067	8.2	103	-0.01
32	Bis(2-ethylhexyl)phthalate	2.865	2.168	24.3	79	0.00
33	Indeno(1,2,3-cd)pyrene	1.249	1.097	12.2	91	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	100	-0.02
35	Benzo(b)fluoranthene	1.151	1.016	11.7	94	-0.01
36	Benzo(k)fluoranthene	1.199	1.074	10.4	99	-0.01
37 C	Benzo(a)pyrene	1.122	0.996	11.2	94	-0.01
38	Dibenzo(a,h)anthracene	1.099	0.944	14.1	91	-0.02
39	Benzo(a,h,i)perylene	1.126	0.958	14.9	93	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101618\
Data File : BE097948.D
Acq On : 17 Oct 2018 15:49
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 17 16:20:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			