


Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101618\
 Data File : BE097931.D
 Acq On : 17 Oct 2018 4:43
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101618.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Oct 16 15:36:44 2018



AutoFind: Scans 2579, 2580, 2581; Background Corrected with Scan 2572

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.0	22605	PASS
68	69	0.00	2	0.7	139	PASS
69	198	0.00	100	43.7	20172	PASS
70	69	0.00	2	0.5	102	PASS
127	198	10	80	44.3	20425	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	46132	PASS
199	198	5	9	6.8	3126	PASS
275	198	10	60	27.9	12883	PASS
365	198	1	100	3.9	1803	PASS
441	198	0.01	100	15.1	6969	PASS
442	198	50	100	92.9	42877	PASS
443	442	15	24	19.5	8359	PASS