

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101718\
 Data File : BE097950.D
 Acq On : 17 Oct 2018 17:02
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 18 10:14:15 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	0.00
2	1,4-Dioxane	0.400	0.414	-3.5	98	0.00
3	n-Nitrosodimethylamine	0.400	0.330	17.5	82	0.00
4 S	2-Fluorophenol	0.400	0.375	6.3	87	0.00
5 S	Phenol-d6	0.400	0.360	10.0	81	0.00
6	bis(2-Chloroethyl)ether	0.400	0.357	10.8	84	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	88	-0.01
8 S	Nitrobenzene-d5	0.400	0.364	9.0	82	-0.01
9	Naphthalene	0.400	0.382	4.5	86	-0.01
10	Hexachlorobutadiene	0.400	0.377	5.8	86	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.365	8.8	83	-0.01
12	2-Methylnaphthalene	0.400	0.364	9.0	83	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	90	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.304	24.0	77	0.00
15 S	2-Fluorobiphenyl	0.400	0.351	12.3	84	0.00
16	Acenaphthylene	0.400	0.346	13.5	82	0.00
17	Acenaphthene	0.400	0.361	9.8	86	0.00
18	Fluorene	0.400	0.359	10.3	85	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	-0.01
20	4-Bromophenyl-phenylether	0.400	0.350	12.5	86	0.00
21	Hexachlorobenzene	0.400	0.360	10.0	88	0.00
22	Pentachlorophenol	0.400	0.346	13.5	89	-0.01
23	Phenanthrene	0.400	0.362	9.5	90	0.00
24	Anthracene	0.400	0.344	14.0	86	0.00
25 SURR	Fluoranthene-d10	0.400	0.352	12.0	90	0.00
26	Fluoranthene	0.400	0.358	10.5	92	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	101	-0.01
28	Pvrene	0.400	0.343	14.2	92	0.00
29 S	Terphenyl-d14	0.400	0.329	17.8	92	0.00
30	Benzo(a)anthracene	0.400	0.348	13.0	93	0.00
31	Chrysene	0.400	0.370	7.5	102	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.257	35.8#	77	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.330	17.5	92	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	100	-0.01
35	Benzo(b)fluoranthene	0.400	0.356	11.0	95	0.00
36	Benzo(k)fluoranthene	0.400	0.356	11.0	98	0.00
37 C	Benzo(a)pyrene	0.400	0.353	11.8	93	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.342	14.5	91	-0.02
39	Benzo(a,h,i)perylene	0.400	0.342	14.5	93	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101718\
Data File : BE097950.D
Acq On : 17 Oct 2018 17:02
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 18 10:14:15 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				