

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101718\
 Data File : BE097950.D
 Acq On : 17 Oct 2018 17:02
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 18 10:14:15 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.697	0.721	-3.4	98	0.00
3	n-Nitrosodimethylamine	0.779	0.642	17.6	82	0.00
4 S	2-Fluorophenol	1.261	1.184	6.1	87	0.00
5 S	Phenol-d6	1.864	1.680	9.9	81	0.00
6	bis(2-Chloroethyl)ether	1.583	1.413	10.7	84	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	88	-0.01
8 S	Nitrobenzene-d5	0.379	0.344	9.2	82	-0.01
9	Naphthalene	3.983	3.799	4.6	86	-0.01
10	Hexachlorobutadiene	0.215	0.202	6.0	86	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.607	8.7	83	-0.01
12	2-Methylnaphthalene	0.682	0.621	8.9	83	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	90	-0.01
14 S	2,4,6-Tribromophenol	0.215	0.163	24.2	77	0.00
15 S	2-Fluorobiphenyl	1.635	1.436	12.2	84	0.00
16	Acenaphthylene	1.848	1.599	13.5	82	0.00
17	Acenaphthene	1.129	1.019	9.7	86	0.00
18	Fluorene	1.554	1.396	10.2	85	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	-0.01
20	4-Bromophenyl-phenylether	0.219	0.191	12.8	86	0.00
21	Hexachlorobenzene	0.215	0.194	9.8	88	0.00
22	Pentachlorophenol	0.062	0.044	29.0#	89	-0.01
23	Phenanthrene	1.016	0.920	9.4	90	0.00
24	Anthracene	0.960	0.826	14.0	86	0.00
25 SURR	Fluoranthene-d10	5.421	4.774	11.9	90	0.00
26	Fluoranthene	1.334	1.194	10.5	92	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	-0.01
28	Pvrene	1.154	0.989	14.3	92	0.00
29 S	Terphenyl-d14	0.917	0.755	17.7	92	0.00
30	Benzo(a)anthracene	1.227	1.067	13.0	93	0.00
31	Chrysene	1.162	1.076	7.4	102	-0.01
32	Bis(2-ethylhexyl)phthalate	2.865	2.120	26.0#	77	0.00
33	Indeno(1,2,3-cd)pyrene	1.249	1.124	10.0	92	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	100	-0.01
35	Benzo(b)fluoranthene	1.151	1.024	11.0	95	0.00
36	Benzo(k)fluoranthene	1.199	1.068	10.9	98	0.00
37 C	Benzo(a)pyrene	1.122	0.991	11.7	93	-0.01
38	Dibenzo(a,h)anthracene	1.099	0.939	14.6	91	-0.02
39	Benzo(a,h,i)perylene	1.126	0.963	14.5	93	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101718\
Data File : BE097950.D
Acq On : 17 Oct 2018 17:02
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 18 10:14:15 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			