

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101718\
 Data File : BE097966.D
 Acq On : 18 Oct 2018 2:53
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 18 10:19:56 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.697	0.702	-0.7	125	0.00
3	n-Nitrosodimethylamine	0.779	0.805	-3.3	134	0.00
4 S	2-Fluorophenol	1.261	1.324	-5.0	127	-0.01
5 S	Phenol-d6	1.864	2.065	-10.8	131	-0.01
6	bis(2-Chloroethyl)ether	1.583	1.639	-3.5	127	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	127	-0.01
8 S	Nitrobenzene-d5	0.379	0.387	-2.1	133	-0.01
9	Naphthalene	3.983	4.343	-9.0	142	-0.01
10	Hexachlorobutadiene	0.215	0.231	-7.4	142	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.747	-12.3	147	-0.01
12	2-Methylnaphthalene	0.682	0.765	-12.2	148	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	134	-0.01
14 S	2,4,6-Tribromophenol	0.215	0.218	-1.4	154#	-0.01
15 S	2-Fluorobiphenyl	1.635	1.634	0.1	142	-0.01
16	Acenaphthylene	1.848	1.977	-7.0	151#	0.00
17	Acenaphthene	1.129	1.197	-6.0	150	0.00
18	Fluorene	1.554	1.599	-2.9	145	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	128	-0.01
20	4-Bromophenyl-phenylether	0.219	0.238	-8.7	143	0.00
21	Hexachlorobenzene	0.215	0.244	-13.5	149	0.00
22	Pentachlorophenol	0.062	0.088	-41.9#	240#	-0.01
23	Phenanthrene	1.016	1.075	-5.8	140	0.00
24	Anthracene	0.960	1.030	-7.3	144	-0.01
25 SURR	Fluoranthene-d10	5.421	5.488	-1.2	139	0.00
26	Fluoranthene	1.334	1.353	-1.4	139	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	119	-0.01
28	Pvrene	1.154	1.254	-8.7	138	0.00
29 S	Terphenyl-d14	0.917	0.959	-4.6	138	-0.01
30	Benzo(a)anthracene	1.227	1.299	-5.9	134	0.00
31	Chrysene	1.162	1.265	-8.9	142	-0.01
32	Bis(2-ethylhexyl)phthalate	2.865	3.139	-9.6	134	0.00
33	Indeno(1,2,3-cd)pyrene	1.249	1.332	-6.6	128	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	118	-0.01
35	Benzo(b)fluoranthene	1.151	1.212	-5.3	133	-0.01
36	Benzo(k)fluoranthene	1.199	1.254	-4.6	136	0.00
37 C	Benzo(a)pyrene	1.122	1.194	-6.4	133	-0.01
38	Dibenzo(a,h)anthracene	1.099	1.146	-4.3	131	-0.02
39	Benzo(a,h,i)perylene	1.126	1.135	-0.8	130	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101718\
Data File : BE097966.D
Acq On : 18 Oct 2018 2:53
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 18 10:19:56 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			