

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE101818\
 Data File : BE097981.D
 Acq On : 18 Oct 2018 17:24
 Operator : SJ/JU
 Sample : J5364-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 KY038PS035-181008

Quant Time: Oct 19 01:51:56 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	2288	0.40	ng	0.00
7) Naphthalene-d8	10.68	136	10920	0.40	ng	-0.01
13) Acenaphthene-d10	14.54	164	8399	0.40	ng	-0.01
19) Phenanthrene-d10	17.29	188	20825	0.40	ng	-0.01
27) Chrysene-d12	21.48	240	22481	0.40	ng	-0.01
34) Perylene-d12	24.03	264	22393	0.40	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.47	112	1367	0.19	ng	0.00
5) Phenol-d6	7.04	99	1256	0.12	ng	-0.01
8) Nitrobenzene-d5	9.03	82	4061	0.39	ng	-0.01
11) 2-Methylnaphthalene-d10	12.28	152	7536	0.41	ng	-0.01
14) 2,4,6-Tribromophenol	16.03	330	1926	0.43	ng	-0.01
15) 2-Fluorobiphenyl	13.16	172	12305	0.36	ng	-0.01
25) Fluoranthene-d10	19.33	212	100282	0.36	ng	0.00
29) Terphenyl-d14	19.92	244	27762	0.54	ng	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE101818\
Data File : BE097981.D
Acq On : 18 Oct 2018 17:24
Operator : SJ/JU
Sample : J5364-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
KY038PS035-181008

Quant Time: Oct 19 01:51:56 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

