

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE101818\
 Data File : BE097968.D
 Acq On : 18 Oct 2018 9:18
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 01:15:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.697	0.754	-8.2	124	0.00
3	n-Nitrosodimethylamine	0.779	0.767	1.5	118	0.00
4 S	2-Fluorophenol	1.261	1.376	-9.1	122	-0.01
5 S	Phenol-d6	1.864	2.233	-19.8	130	-0.01
6	bis(2-Chloroethyl)ether	1.583	1.716	-8.4	122	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	121	-0.01
8 S	Nitrobenzene-d5	0.379	0.423	-11.6	137	-0.01
9	Naphthalene	3.983	4.471	-12.3	139	-0.01
10	Hexachlorobutadiene	0.215	0.241	-12.1	140	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.735	-10.5	137	-0.01
12	2-Methylnaphthalene	0.682	0.780	-14.4	143	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	127	-0.01
14 S	2,4,6-Tribromophenol	0.215	0.217	-0.9	145	0.00
15 S	2-Fluorobiphenyl	1.635	1.712	-4.7	141	-0.01
16	Acenaphthylene	1.848	2.024	-9.5	147	0.00
17	Acenaphthene	1.129	1.227	-8.7	146	0.00
18	Fluorene	1.554	1.651	-6.2	142	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	124	-0.01
20	4-Bromophenyl-phenylether	0.219	0.240	-9.6	139	0.00
21	Hexachlorobenzene	0.215	0.250	-16.3	147	0.00
22	Pentachlorophenol	0.062	0.081	-30.6#	214#	-0.01
23	Phenanthrene	1.016	1.080	-6.3	136	0.00
24	Anthracene	0.960	1.017	-5.9	137	0.00
25 SURR	Fluoranthene-d10	5.421	5.274	2.7	129	0.00
26	Fluoranthene	1.334	1.332	0.1	133	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	113	-0.01
28	Pyrene	1.154	1.277	-10.7	134	0.00
29 S	Terphenyl-d14	0.917	0.994	-8.4	136	-0.01
30	Benzo(a)anthracene	1.227	1.339	-9.1	131	0.00
31	Chrysene	1.162	1.292	-11.2	138	-0.01
32	Bis(2-ethylhexyl)phthalate	2.865	3.011	-5.1	122	0.00
33	Indeno(1,2,3-cd)pyrene	1.249	1.375	-10.1	126	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	112	-0.01
35	Benzo(b)fluoranthene	1.151	1.233	-7.1	129	-0.01
36	Benzo(k)fluoranthene	1.199	1.288	-7.4	133	-0.01
37 C	Benzo(a)pyrene	1.122	1.215	-8.3	129	-0.01
38	Dibenzo(a,h)anthracene	1.099	1.166	-6.1	127	-0.01
39	Benzo(g,h,i)perylene	1.126	1.171	-4.0	128	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE101818\
Data File : BE097968.D
Acq On : 18 Oct 2018 9:18
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 01:15:19 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		