

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE101818\
 Data File : BE097983.D
 Acq On : 18 Oct 2018 18:36
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 01:55:54 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	-0.01
2	1,4-Dioxane	0.697	0.644	7.6	119	0.00
3	n-Nitrosodimethylamine	0.779	0.689	11.6	119	0.00
4 S	2-Fluorophenol	1.261	1.321	-4.8	132	-0.01
5 S	Phenol-d6	1.864	2.119	-13.7	139	-0.01
6	bis(2-Chloroethyl)ether	1.583	1.581	0.1	127	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	134	-0.01
8 S	Nitrobenzene-d5	0.379	0.393	-3.7	142	-0.01
9	Naphthalene	3.983	4.275	-7.3	147	-0.01
10	Hexachlorobutadiene	0.215	0.240	-11.6	154#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.731	-9.9	150#	-0.02
12	2-Methylnaphthalene	0.682	0.759	-11.3	154#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	143	-0.02
14 S	2,4,6-Tribromophenol	0.215	0.236	-9.8	177#	-0.01
15 S	2-Fluorobiphenyl	1.635	1.603	2.0	148	-0.02
16	Acenaphthylene	1.848	1.982	-7.3	161#	0.00
17	Acenaphthene	1.129	1.198	-6.1	159#	0.00
18	Fluorene	1.554	1.598	-2.8	154#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	136	-0.01
20	4-Bromophenyl-phenylether	0.219	0.244	-11.4	155#	0.00
21	Hexachlorobenzene	0.215	0.253	-17.7	163#	0.00
22	Pentachlorophenol	0.062	0.098	-58.1#	284#	-0.01
23	Phenanthrene	1.016	1.064	-4.7	147	0.00
24	Anthracene	0.960	1.030	-7.3	152#	-0.01
25 SURR	Fluoranthene-d10	5.421	5.490	-1.3	147	-0.01
26	Fluoranthene	1.334	1.339	-0.4	146	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	124	-0.01
28	Pyrene	1.154	1.267	-9.8	145	-0.01
29 S	Terphenyl-d14	0.917	0.991	-8.1	147	-0.01
30	Benzo(a)anthracene	1.227	1.270	-3.5	136	0.00
31	Chrysene	1.162	1.247	-7.3	145	-0.01
32	Bis(2-ethylhexyl)phthalate	2.865	3.410	-19.0	151#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.249	1.344	-7.6	134	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	122	-0.01
35	Benzo(b)fluoranthene	1.151	1.201	-4.3	137	-0.01
36	Benzo(k)fluoranthene	1.199	1.235	-3.0	139	-0.01
37 C	Benzo(a)pyrene	1.122	1.168	-4.1	135	-0.01
38	Dibenzo(a,h)anthracene	1.099	1.139	-3.6	135	-0.02
39	Benzo(g,h,i)perylene	1.126	1.136	-0.9	135	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE101818\
Data File : BE097983.D
Acq On : 18 Oct 2018 18:36
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 01:55:54 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			