

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101819\
 Data File : BE100660.D
 Acq On : 18 Oct 2019 19:06
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 00:00:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 10 14:01:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.714	0.567	20.6	86	0.03
3	n-Nitrosodimethylamine	0.755	0.694	8.1	107	0.02
4 S	2-Fluorophenol	1.235	1.351	-9.4	124	0.02
5 S	Phenol-d6	1.742	1.959	-12.5	127	0.01
6	bis(2-Chloroethyl)ether	1.430	1.410	1.4	113	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.248	0.381	-53.6#	186#	0.00
9	Naphthalene	4.206	4.117	2.1	107	-0.01
10	Hexachlorobutadiene	0.183	0.189	-3.3	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.631	0.599	5.1	105	0.00
12	2-Methylnaphthalene	0.717	0.685	4.5	106	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	-0.01
14 S	2,4,6-Tribromophenol	0.106	0.162	-52.8#	168#	0.00
15 S	2-Fluorobiphenyl	1.457	1.651	-13.3	112	-0.01
16	Acenaphthylene	1.721	1.921	-11.6	112	0.00
17	Acenaphthene	1.112	1.084	2.5	98	0.00
18	Fluorene	1.477	1.538	-4.1	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.204	0.211	-3.4	94	-0.01
21	Hexachlorobenzene	0.194	0.194	0.0	92	0.00
22	Pentachlorophenol	0.062	0.107	-72.6#	169#	0.00
23	Phenanthrene	1.100	1.107	-0.6	87	0.00
24	Anthracene	1.021	1.042	-2.1	93	0.00
25 SURR	Fluoranthene-d10	5.162	4.602	10.8	81	0.00
26	Fluoranthene	1.436	1.336	7.0	82	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	68	0.00
28	Pvrene	1.118	1.348	-20.6	82	0.00
29 S	Terphenyl-d14	0.875	1.078	-23.2	88	0.00
30	Benzo(a)anthracene	1.305	1.363	-4.4	72	0.00
31	Chrysene	1.304	1.307	-0.2	69	0.00
32	Bis(2-ethylhexyl)phthalate	2.729	3.585	-31.4#	89	-0.01
33	Indeno(1,2,3-cd)pyrene	1.603	1.533	4.4	67	0.00
34 I	Perylene-d12	1.000	1.000	0.0	61	0.00
35	Benzo(b)fluoranthene	1.166	1.170	-0.3	66	0.00
36	Benzo(k)fluoranthene	1.200	1.168	2.7	63	0.00
37 C	Benzo(a)pyrene	1.086	1.111	-2.3	68	0.00
38	Dibenzo(a,h)anthracene	1.122	1.106	1.4	67	0.00
39	Benzo(a,h,i)perylene	1.129	1.138	-0.8	66	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE101819\
Data File : BE100660.D
Acq On : 18 Oct 2019 19:06
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 00:00:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 10 14:01:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			