

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_E\DATA\BE102218\  
 Data File : BE098028.D  
 Acq On : 23 Oct 2018 2:41  
 Operator : SJ/JU  
 Sample : J5551-02  
 Misc :  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 C0TY2

Manual Integrations  
 APPROVED

Sohil  
 10/23/2018 5:35:09 PM

Quant Time: Oct 23 04:26:40 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\SOM-EPA-SIM-BE101018.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Oct 23 02:34:53 2018  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.88  | 152  | 1634     | 0.40 | ng/ul | 0.00     |
| 2) Naphthalene-d8         | 10.68 | 136  | 7900     | 0.40 | ng/ul | 0.00     |
| 6) Acenaphthene-d10       | 14.54 | 164  | 5986     | 0.40 | ng/ul | 0.00     |
| 10) Phenanthrene-d10      | 17.30 | 188  | 13284m   | 0.40 | ng/ul | 0.00     |
| 16) Chrysene-d12          | 21.48 | 240  | 14806m   | 0.40 | ng/ul | 0.00     |
| 20) Perylene-d12          | 24.03 | 264  | 13682m   | 0.40 | ng/ul | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|-------|------|----------|------|-------|----------|
| 4) 2-Methylnaphthalene-d10  | 12.28 | 152  | 4961     | 0.40 | ng/ul | 0.00     |
| 14) Fluoranthene-d10        | 19.33 | 212  | 15834m   | 0.37 | ng/ul | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_E\DATA\BE102218\  
Data File : BE098028.D  
Acq On : 23 Oct 2018 2:41  
Operator : SJ/JU  
Sample : J5551-02  
Misc :  
ALS Vial : 26 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
C0TY2

Quant Time: Oct 23 04:26:40 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\SOM-EPA-SIM-BE101018.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Oct 23 02:34:53 2018  
Response via : Initial Calibration

Manual Integrations  
APPROVED

Sohil  
10/23/2018 5:35:09 PM

