

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE102319\
 Data File : BE100673.D
 Acq On : 23 Oct 2019 22:09
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 24 01:34:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 16:49:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	59	0.00
2	1,4-Dioxane	0.714	0.624	12.6	52	0.00
3	n-Nitrosodimethylamine	0.755	0.628	16.8	53	0.00
4 S	2-Fluorophenol	1.235	1.236	-0.1	62	0.00
5 S	Phenol-d6	1.742	1.772	-1.7	63	0.00
6	bis(2-Chloroethyl)ether	1.430	1.293	9.6	57	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	60	0.00
8 S	Nitrobenzene-d5	0.248	0.327	-31.9#	91	0.00
9	Naphthalene	4.206	3.973	5.5	58	0.00
10	Hexachlorobutadiene	0.183	0.177	3.3	61	0.00
11 SURR	2-Methylnaphthalene-d10	0.631	0.593	6.0	59	0.00
12	2-Methylnaphthalene	0.717	0.665	7.3	58	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	63	0.00
14 S	2,4,6-Tribromophenol	0.106	0.140	-32.1#	94	0.00
15 S	2-Fluorobiphenyl	1.457	1.504	-3.2	66	0.00
16	Acenaphthylene	1.721	1.547	10.1	59	-0.01
17	Acenaphthene	1.112	1.017	8.5	60	0.00
18	Fluorene	1.477	1.353	8.4	60	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	67	-0.01
20	4-Bromophenyl-phenylether	0.204	0.186	8.8	63	0.00
21	Hexachlorobenzene	0.194	0.182	6.2	66	-0.01
22	Pentachlorophenol	0.062	0.078	-25.8#	94	0.00
23	Phenanthrene	1.100	1.061	3.5	64	-0.01
24	Anthracene	1.021	0.872	14.6	59	0.00
25 SURR	Fluoranthene-d10	5.162	4.848	6.1	65	0.00
26	Fluoranthene	1.436	1.420	1.1	66	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	70	0.00
28	Pvrene	1.118	1.062	5.0	67	0.00
29 S	Terphenyl-d14	0.875	0.903	-3.2	76	0.00
30	Benzo(a)anthracene	1.305	1.181	9.5	64	0.00
31	Chrysene	1.304	1.245	4.5	67	0.00
32	Bis(2-ethylhexyl)phthalate	2.729	2.466	9.6	63	-0.01
33	Indeno(1,2,3-cd)pyrene	1.603	1.443	10.0	65	0.00
34 I	Perylene-d12	1.000	1.000	0.0	68	0.00
35	Benzo(b)fluoranthene	1.166	1.032	11.5	65	0.00
36	Benzo(k)fluoranthene	1.200	1.093	8.9	66	0.00
37 C	Benzo(a)pyrene	1.086	0.952	12.3	64	0.00
38	Dibenzo(a,h)anthracene	1.122	0.981	12.6	65	0.00
39	Benzo(a,h,i)perylene	1.129	1.010	10.5	65	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE102319\
Data File : BE100673.D
Acq On : 23 Oct 2019 22:09
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 24 01:34:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 16:49:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			