

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE102418\
 Data File : BE098038.D
 Acq On : 24 Oct 2018 14:45
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICV0.4

Quant Time: Oct 24 15:16:54 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE102418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 24 13:49:35 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	94	0.00
2	1,4-Dioxane	0.400	0.386	3.5	97	0.00
3	n-Nitrosodimethylamine	0.400	0.401	-0.3	100	0.00
4 S	2-Fluorophenol	0.400	0.399	0.3	97	0.00
5 S	Phenol-d6	0.400	0.389	2.8	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.380	5.0	92	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	94	0.01
8 S	Nitrobenzene-d5	0.400	0.400	0.0	98	0.01
9	Naphthalene	0.400	0.390	2.5	95	0.01
10	Hexachlorobutadiene	0.400	0.394	1.5	98	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	97	0.01
12	2-Methylnaphthalene	0.400	0.392	2.0	94	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	0.01
14 S	2,4,6-Tribromophenol	0.400	0.369	7.8	95	0.01
15 S	2-Fluorobiphenyl	0.400	0.373	6.8	95	0.01
16	Acenaphthylene	0.400	0.385	3.8	97	0.00
17	Acenaphthene	0.400	0.393	1.8	97	0.01
18	Fluorene	0.400	0.380	5.0	95	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.400	0.397	0.8	96	0.01
21	Hexachlorobenzene	0.400	0.395	1.3	98	0.00
22	Pentachlorophenol	0.400	0.429	-7.2	114	0.01
23	Phenanthrene	0.400	0.390	2.5	97	0.00
24	Anthracene	0.400	0.385	3.8	97	0.01
25 SURR	Fluoranthene-d10	0.400	0.383	4.3	95	0.00
26	Fluoranthene	0.400	0.383	4.3	95	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	94	0.00
28	Pvrene	0.400	0.377	5.8	96	0.00
29 S	Terphenyl-d14	0.400	0.370	7.5	97	0.00
30	Benzo(a)anthracene	0.400	0.381	4.8	97	0.00
31	Chrysene	0.400	0.380	5.0	96	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.378	5.5	96	0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	96	0.00
34 I	Perylene-d12	0.400	0.400	0.0	97	0.00
35	Benzo(b)fluoranthene	0.400	0.376	6.0	96	0.00
36	Benzo(k)fluoranthene	0.400	0.367	8.3	95	0.00
37 C	Benzo(a)pyrene	0.400	0.376	6.0	98	0.00
38	Dibenzo(a,h)anthracene	0.400	0.370	7.5	95	0.01
39	Benzo(a,h,i)perylene	0.400	0.376	6.0	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE102418\
Data File : BE098038.D
Acq On : 24 Oct 2018 14:45
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICV0.4

Quant Time: Oct 24 15:16:54 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE102418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 24 13:49:35 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				