

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE102518\
 Data File : BE098057.D
 Acq On : 25 Oct 2018 16:07
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.468

Quant Time: Oct 25 16:43:46 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE102518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 25 14:48:33 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
3	Naphthalene	0.957	0.914	4.5	97	0.00
4 SURR	2-Methylnaphthalene-d10	0.713	0.685	3.9	99	0.00
5	2-Methylnaphthalene	0.724	0.692	4.4	97	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
7	Acenaphthylene	1.578	1.464	7.2	95	0.00
8 C	Acenaphthene	1.179	1.086	7.9	95	0.00
9	Fluorene	1.419	1.326	6.6	96	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
11	Pentachlorophenol	0.021	0.018	14.3	87	0.00
12	Phenanthrene	0.923	0.869	5.9	96	0.00
13	Anthracene	0.912	0.853	6.5	96	0.00
14 SURR	Fluoranthene-d10	1.300	1.221	6.1	95	0.00
15 C	Fluoranthene	1.177	1.101	6.5	95	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
17	Pyrene	0.934	0.956	-2.4	97	0.00
18	Benzo(a)anthracene	1.176	0.967	17.8	77	0.00
19	Chrysene	0.910	0.895	1.6	95	0.00
20 I	Perylene-d12	1.000	1.000	0.0	101	0.00
21	Benzo(b)fluoranthene	1.108	1.063	4.1	97	0.00
22	Benzo(k)fluoranthene	1.085	0.988	8.9	94	0.00
23 C	Benzo(a)pyrene	1.066	0.985	7.6	96	0.00
24	Indeno(1,2,3-cd)pyrene	1.221	1.152	5.7	98	-0.01
25	Dibenzo(a,h)anthracene	1.018	0.959	5.8	97	0.00
26	Benzo(g,h,i)perylene	1.028	0.965	6.1	96	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0