

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE102518\
 Data File : BE098079.D
 Acq On : 26 Oct 2018 5:41
 Operator : SJ/JU
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.470

Quant Time: Oct 26 07:46:40 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE102518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 26 07:43:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.88	152	986	0.40	ng/ul	0.00
2) Naphthalene-d8	10.68	136	4849	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.53	164	3353	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.30	188	8178	0.40	ng/ul	0.00
16) Chrysene-d12	21.47	240	12161	0.40	ng/ul	0.00
20) Perylene-d12	24.03	264	10149	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.27	152	3226	0.37	ng/ul	0.00
14) Fluoranthene-d10	19.32	212	10161	0.38	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	10.73	128	4399	0.379	ng/ul	100
5) 2-Methylnaphthalene	12.35	142	3379	0.385	ng/ul	100
7) Acenaphthylene	14.25	152	5027	0.380	ng/ul	98
8) Acenaphthene	14.60	153	3708	0.375	ng/ul	99
9) Fluorene	15.59	166	4473	0.376	ng/ul	99
11) Pentachlorophenol	16.96	266	152	0.355	ng/ul#	85
12) Phenanthrene	17.33	178	7096	0.376	ng/ul	98
13) Anthracene	17.42	178	7142	0.383	ng/ul	99
15) Fluoranthene	19.35	202	9173	0.381	ng/ul	99
17) Pyrene	19.71	202	9270	0.326	ng/ul	98
18) Benzo(a)anthracene	21.45	228	12918	0.361	ng/ul	99
19) Chrysene	21.51	228	10280	0.372	ng/ul	100
21) Benzo(b)fluoranthene	23.24	252	10607	0.377	ng/ul	94
22) Benzo(k)fluoranthene	23.29	252	10359	0.376	ng/ul#	95
23) Benzo(a)pyrene	23.91	252	10253	0.379	ng/ul	97
24) Indeno(1,2,3-cd)pyrene	26.72	276	11973	0.386	ng/ul#	98
25) Dibenzo(a,h)anthracene	26.74	278	10008	0.387	ng/ul#	94
26) Benzo(g,h,i)perylene	27.55	276	9951	0.381	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE102518\
Data File : BE098079.D
Acq On : 26 Oct 2018 5:41
Operator : SJ/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.470

Quant Time: Oct 26 07:46:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE102518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 26 07:43:08 2018
Response via : Initial Calibration

