

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE102919\
 Data File : BE100694.D
 Acq On : 29 Oct 2019 14:26
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 29 16:45:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 16:49:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	67	0.00
2	1,4-Dioxane	0.714	0.782	-9.5	74	0.00
3	n-Nitrosodimethylamine	0.755	0.776	-2.8	75	0.00
4 S	2-Fluorophenol	1.235	1.287	-4.2	74	0.00
5 S	Phenol-d6	1.742	1.918	-10.1	78	0.00
6	bis(2-Chloroethyl)ether	1.430	1.625	-13.6	81	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
8 S	Nitrobenzene-d5	0.248	0.321	-29.4#	106	0.00
9	Naphthalene	4.206	4.669	-11.0	82	0.00
10	Hexachlorobutadiene	0.183	0.208	-13.7	85	0.00
11 SURR	2-Methylnaphthalene-d10	0.631	0.704	-11.6	84	0.00
12	2-Methylnaphthalene	0.717	0.808	-12.7	84	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	73	0.00
14 S	2,4,6-Tribromophenol	0.106	0.133	-25.5#	105	0.00
15 S	2-Fluorobiphenyl	1.457	1.586	-8.9	82	0.00
16	Acenaphthylene	1.721	1.734	-0.8	77	-0.01
17	Acenaphthene	1.112	1.199	-7.8	82	0.00
18	Fluorene	1.477	1.676	-13.5	86	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	78	-0.01
20	4-Bromophenyl-phenylether	0.204	0.221	-8.3	88	0.00
21	Hexachlorobenzene	0.194	0.215	-10.8	91	-0.01
22	Pentachlorophenol	0.062	0.052	16.1	73	0.00
23	Phenanthrene	1.100	1.239	-12.6	88	-0.01
24	Anthracene	1.021	1.058	-3.6	84	0.00
25 SURR	Fluoranthene-d10	5.162	5.620	-8.9	89	0.00
26	Fluoranthene	1.436	1.599	-11.4	87	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	69	0.00
28	Pvrene	1.118	1.355	-21.2	85	0.00
29 S	Terphenyl-d14	0.875	1.026	-17.3	86	0.00
30	Benzo(a)anthracene	1.305	1.417	-8.6	76	0.00
31	Chrysene	1.304	1.449	-11.1	78	0.00
32	Bis(2-ethylhexyl)phthalate	2.729	2.367	13.3	60	-0.01
33	Indeno(1,2,3-cd)pyrene	1.603	1.705	-6.4	76	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	66	-0.01
35	Benzo(b)fluoranthene	1.166	1.206	-3.4	73	0.00
36	Benzo(k)fluoranthene	1.200	1.337	-11.4	78	-0.01
37 C	Benzo(a)pyrene	1.086	1.103	-1.6	72	-0.01
38	Dibenzo(a,h)anthracene	1.122	1.164	-3.7	75	-0.01
39	Benzo(a,h,i)perylene	1.129	1.214	-7.5	76	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE102919\
Data File : BE100694.D
Acq On : 29 Oct 2019 14:26
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 29 16:45:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 16:49:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			