

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE102919\
 Data File : BE100698.D
 Acq On : 29 Oct 2019 19:03
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 29 19:46:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 16:49:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	71	0.00
2	1,4-Dioxane	0.400	0.433	-8.2	78	0.00
3	n-Nitrosodimethylamine	0.400	0.416	-4.0	80	0.00
4 S	2-Fluorophenol	0.400	0.428	-7.0	80	0.00
5 S	Phenol-d6	0.400	0.456	-14.0	86	0.00
6	bis(2-Chloroethyl)ether	0.400	0.464	-16.0	88	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	79	0.00
8 S	Nitrobenzene-d5	0.400	0.540	-35.0#	121	0.00
9	Naphthalene	0.400	0.446	-11.5	90	0.00
10	Hexachlorobutadiene	0.400	0.444	-11.0	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.446	-11.5	91	0.00
12	2-Methylnaphthalene	0.400	0.450	-12.5	92	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	81	0.00
14 S	2,4,6-Tribromophenol	0.400	0.544	-36.0#	126	0.00
15 S	2-Fluorobiphenyl	0.400	0.428	-7.0	89	0.00
16	Acenaphthylene	0.400	0.409	-2.2	86	-0.01
17	Acenaphthene	0.400	0.444	-11.0	94	0.00
18	Fluorene	0.400	0.447	-11.7	94	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	88	-0.01
20	4-Bromophenyl-phenylether	0.400	0.439	-9.7	100	0.00
21	Hexachlorobenzene	0.400	0.431	-7.7	100	-0.01
22	Pentachlorophenol	0.400	0.350	12.5	87	0.00
23	Phenanthrene	0.400	0.450	-12.5	98	-0.01
24	Anthracene	0.400	0.431	-7.7	99	0.00
25 SURR	Fluoranthene-d10	0.400	0.431	-7.7	99	0.00
26	Fluoranthene	0.400	0.437	-9.2	97	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	76	0.00
28	Pvrene	0.400	0.494	-23.5	94	0.00
29 S	Terphenyl-d14	0.400	0.472	-18.0	95	0.00
30	Benzo(a)anthracene	0.400	0.446	-11.5	86	-0.01
31	Chrysene	0.400	0.443	-10.7	85	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.406	-1.5	77	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.425	-6.2	84	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	72	-0.01
35	Benzo(b)fluoranthene	0.400	0.420	-5.0	82	0.00
36	Benzo(k)fluoranthene	0.400	0.441	-10.2	85	-0.01
37 C	Benzo(a)pyrene	0.400	0.413	-3.2	80	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.426	-6.5	85	-0.01
39	Benzo(a,h,i)perylene	0.400	0.432	-8.0	83	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE102919\
Data File : BE100698.D
Acq On : 29 Oct 2019 19:03
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 29 19:46:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 16:49:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				