

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE102924\
 Data File : BE101414.D
 Acq On : 29 Oct 2024 19:16
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 30 00:27:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE102824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 29 01:26:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	39.290	1.8	103	0.00
3	Pyridine	40.000	35.303	11.7	92	0.00
4	n-Nitrosodimethylamine	40.000	41.020	-2.6	103	0.00
5 S	2-Fluorophenol	80.000	79.479	0.7	102	0.00
6	Aniline	40.000	51.382	-28.5#	120	0.00
7 S	Phenol-d6	80.000	82.718	-3.4	104	0.00
8	2-Chlorophenol	40.000	40.518	-1.3	102	0.00
9	Benzaldehyde	40.000	49.914	-24.8	122	0.00
10 C	Phenol	40.000	42.453	-6.1	103	0.00
11	bis(2-Chloroethyl)ether	40.000	35.529	11.2	88	0.00
12	1,3-Dichlorobenzene	40.000	39.545	1.1	101	0.00
13 C	1,4-Dichlorobenzene	40.000	39.169	2.1	101	0.00
14	1,2-Dichlorobenzene	40.000	40.007	-0.0	102	0.00
15	Benzyl Alcohol	40.000	44.776	-11.9	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.157	-2.9	105	0.00
17	2-Methylphenol	40.000	41.628	-4.1	103	0.00
18	Hexachloroethane	40.000	40.950	-2.4	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	45.227	-13.1	107	0.00
20	3+4-Methylphenols	40.000	42.504	-6.3	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	108	0.00
22	Acetophenone	40.000	40.957	-2.4	105	0.00
23 S	Nitrobenzene-d5	80.000	82.269	-2.8	106	0.00
24	Nitrobenzene	40.000	40.484	-1.2	104	0.00
25	Isophorone	40.000	41.735	-4.3	105	0.00
26 C	2-Nitrophenol	40.000	42.740	-6.9	106	0.00
27	2,4-Dimethylphenol	40.000	40.180	-0.4	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.186	-3.0	107	0.00
29 C	2,4-Dichlorophenol	40.000	41.075	-2.7	106	0.00
30	1,2,4-Trichlorobenzene	40.000	39.373	1.6	105	0.00
31	Naphthalene	40.000	39.771	0.6	105	0.00
32	Benzoic acid	40.000	40.692	-1.7	115	0.01
33	4-Chloroaniline	40.000	43.982	-10.0	106	0.00
34 C	Hexachlorobutadiene	40.000	39.659	0.9	105	0.00
35	Caprolactam	40.000	42.798	-7.0	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.787	-4.5	108	0.00
37	2-Methylnaphthalene	40.000	40.559	-1.4	107	0.00
38	1-Methylnaphthalene	40.000	40.246	-0.6	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.884	0.3	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.119	-0.3	103	0.00
42 S	2,4,6-Tribromophenol	80.000	83.220	-4.0	109	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.923	-4.8	107	0.00
44	2,4,5-Trichlorophenol	40.000	41.547	-3.9	109	0.00
45 S	2-Fluorobiphenyl	80.000	80.326	-0.4	106	0.00
46	1,1'-Biphenyl	40.000	39.858	0.4	106	0.00
47	2-Chloronaphthalene	40.000	39.803	0.5	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE102924\
 Data File : BE101414.D
 Acq On : 29 Oct 2024 19:16
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 30 00:27:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE102824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 29 01:26:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.375	-8.4	109	0.00
49	Acenaphthylene	40.000	40.664	-1.7	106	0.00
50	Dimethylphthalate	40.000	39.720	0.7	106	0.00
51	2,6-Dinitrotoluene	40.000	41.255	-3.1	106	0.00
52 C	Acenaphthene	40.000	39.900	0.3	106	0.00
53	3-Nitroaniline	40.000	43.407	-8.5	108	0.00
54 P	2,4-Dinitrophenol	40.000	38.104	4.7	107	0.00
55	Dibenzofuran	40.000	39.499	1.3	106	0.00
56 P	4-Nitrophenol	40.000	40.821	-2.1	106	0.00
57	2,4-Dinitrotoluene	40.000	42.223	-5.6	108	0.00
58	Fluorene	40.000	40.225	-0.6	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.886	-2.2	107	0.00
60	Diethylphthalate	40.000	39.636	0.9	104	0.00
61	4-Chlorophenyl-phenylether	40.000	39.810	0.5	106	0.00
62	4-Nitroaniline	40.000	42.440	-6.1	107	0.00
63	Azobenzene	40.000	40.208	-0.5	106	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.788	-4.5	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.829	-2.1	105	0.00
67	4-Bromophenyl-phenylether	40.000	40.676	-1.7	107	0.00
68	Hexachlorobenzene	40.000	40.830	-2.1	107	0.00
69	Atrazine	40.000	39.379	1.6	93	0.00
70 C	Pentachlorophenol	40.000	44.204	-10.5	109	0.00
71	Phenanthrene	40.000	40.218	-0.5	104	0.00
72	Anthracene	40.000	41.003	-2.5	105	0.00
73	Carbazole	40.000	40.831	-2.1	105	0.00
74	Di-n-butylphthalate	40.000	41.909	-4.8	104	0.00
75 C	Fluoranthene	40.000	40.495	-1.2	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzidine	40.000	69.143	-72.9#	138	0.00
78	Pyrene	40.000	41.224	-3.1	105	0.00
79 S	Terphenyl-d14	80.000	73.450	8.2	103	0.00
80	Butylbenzylphthalate	40.000	41.462	-3.7	106	0.00
81	Benzo(a)anthracene	40.000	40.735	-1.8	106	0.00
82	3,3'-Dichlorobenzidine	40.000	43.662	-9.2	109	0.00
83	Chrysene	40.000	40.098	-0.2	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.935	-7.3	107	0.00
85 c	Di-n-octyl phthalate	40.000	42.426	-6.1	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	112	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.795	-2.0	110	0.00
88	Benzo(b)fluoranthene	40.000	39.612	1.0	104	0.00
89	Benzo(k)fluoranthene	40.000	40.797	-2.0	109	0.00
90 C	Benzo(a)pyrene	40.000	40.931	-2.3	108	0.00
91	Dibenzo(a,h)anthracene	40.000	41.201	-3.0	109	0.00
92	Benzo(g,h,i)perylene	40.000	40.529	-1.3	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE102924\
Data File : BE101414.D
Acq On : 29 Oct 2024 19:16
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC040

Quant Time: Oct 30 00:27:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE102824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 29 01:26:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0