

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE103018\
 Data File : BE098111.D
 Acq On : 30 Oct 2018 12:51
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.475

Quant Time: Oct 30 13:22:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE102518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 30 12:56:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.87	152	1037	0.40	ng/ul	0.00
2) Naphthalene-d8	10.67	136	5209	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.53	164	3620	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.29	188	8741	0.40	ng/ul	0.00
16) Chrysene-d12	21.47	240	13771	0.40	ng/ul	0.00
20) Perylene-d12	24.02	264	11119	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.27	152	3485	0.38	ng/ul	0.00
14) Fluoranthene-d10	19.32	212	10551	0.37	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.72	128	4837	0.388	ng/ul	97
5) 2-Methylnaphthalene	12.34	142	3573	0.379	ng/ul	99
7) Acenaphthylene	14.25	152	5404	0.378	ng/ul	99
8) Acenaphthene	14.59	153	4031	0.378	ng/ul	95
9) Fluorene	15.59	166	4817	0.375	ng/ul	97
11) Pentachlorophenol	16.95	266	205	0.448	ng/ul	97
12) Phenanthrene	17.33	178	7652	0.379	ng/ul	100
13) Anthracene	17.42	178	7518	0.377	ng/ul	98
15) Fluoranthene	19.35	202	9689	0.377	ng/ul	97
17) Pyrene	19.71	202	12058	0.375	ng/ul	97
18) Benzo(a)anthracene	21.45	228	12797	0.316	ng/ul	99
19) Chrysene	21.51	228	10640	0.340	ng/ul	100
21) Benzo(b)fluoranthene	23.24	252	10630	0.345	ng/ul	98
22) Benzo(k)fluoranthene	23.29	252	10712	0.355	ng/ul	99
23) Benzo(a)pyrene	23.90	252	10164	0.343	ng/ul	99
24) Indeno(1,2,3-cd)pyrene	26.71	276	11750	0.346	ng/ul#	96
25) Dibenzo(a,h)anthracene	26.74	278	9837	0.348	ng/ul	98
26) Benzo(g,h,i)perylene	27.54	276	9828	0.344	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE103018\
Data File : BE098111.D
Acq On : 30 Oct 2018 12:51
Operator : SJ/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.475

Quant Time: Oct 30 13:22:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE102518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 30 12:56:30 2018
Response via : Initial Calibration

