

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE103118\  
 Data File : BE098127.D  
 Acq On : 31 Oct 2018 22:00  
 Operator : SJ/JU  
 Sample : J5701-16  
 Misc :  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 A40G0

Quant Time: Nov 01 03:52:46 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE102518.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Nov 01 02:11:37 2018  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.87  | 152  | 1327     | 0.40 | ng/ul | 0.00     |
| 2) Naphthalene-d8         | 10.67 | 136  | 5875     | 0.40 | ng/ul | 0.00     |
| 6) Acenaphthene-d10       | 14.53 | 164  | 4803     | 0.40 | ng/ul | 0.00     |
| 10) Phenanthrene-d10      | 17.29 | 188  | 48057    | 0.40 | ng/ul | 0.00     |
| 16) Chrysene-d12          | 21.47 | 240  | 2398     | 0.40 | ng/ul | 0.00     |
| 20) Perylene-d12          | 24.03 | 264  | 9289     | 0.40 | ng/ul | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|-------|------|----------|------|-------|----------|
| 4) 2-Methylnaphthalene-d10  | 12.27 | 152  | 1799     | 0.17 | ng/ul | 0.00     |
| 14) Fluoranthene-d10        | 19.32 | 212  | 4883     | 0.03 | ng/ul | 0.00     |

| Target Compounds | Ovalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE103118\  
Data File : BE098127.D  
Acq On : 31 Oct 2018 22:00  
Operator : SJ/JU  
Sample : J5701-16  
Misc :  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
A40G0

Quant Time: Nov 01 03:52:46 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE102518.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Nov 01 02:11:37 2018  
Response via : Initial Calibration

