

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE110118\
 Data File : BE098148.D
 Acq On : 1 Nov 2018 16:11
 Operator : SJ/JU
 Sample : J5665-08
 Misc :
 ALS Vial : 5 Sample Multiplier: 50

Instrument :
 BNA_E
 ClientSampleId :
 TWP-03

Quant Time: Nov 02 13:12:56 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE110118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 13:00:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.87	152	57789	20.00	ng	0.00
7) Naphthalene-d8	10.67	136	269101	20.00	ng	0.00
13) Acenaphthene-d10	14.54	164	187660	20.00	ng	0.00
19) Phenanthrene-d10	17.28	188	439608	20.00	ng	0.00
27) Chrysene-d12	21.48	240	505188	20.00	ng	0.01
34) Perylene-d12	24.02	264	512144	20.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.45	112	207194	60.40	ng	0.00
5) Phenol-d6	7.04	99	214657	37.99	ng	0.00
8) Nitrobenzene-d5	9.03	82	505272	87.54	ng	0.00
11) 2-Methylnaphthalene-d10	12.28	152	48	0.01	ng	0.02
14) 2,4,6-Tribromophenol	16.03	330	261551	147.17	ng	0.00
15) 2-Fluorobiphenyl	13.16	172	1073120	69.54	ng	0.00
25) Fluoranthene-d10	19.32	212	21	0.00	ng	0.00
29) Terphenyl-d14	19.92	244	1757717	74.81	ng	0.00
Target Compounds						
9) Naphthalene	10.72	128	2605	0.048	ng	# 92
30) Benzo(a)anthracene	21.46	228	1702	0.058	ng	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE110118\
Data File : BE098148.D
Acq On : 1 Nov 2018 16:11
Operator : SJ/JU
Sample : J5665-08
Misc :
ALS Vial : 5 Sample Multiplier: 50

Instrument :
BNA_E
ClientSampleId :
TWP-03

Quant Time: Nov 02 13:12:56 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE110118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 13:00:52 2018
Response via : Initial Calibration

