

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE110118\  
 Data File : BE098143.D  
 Acq On : 1 Nov 2018 13:04  
 Operator : SJ/JU  
 Sample : SSTDICV0.4  
 Misc :  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 ICVBE110118

Quant Time: Nov 01 13:34:12 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE110118.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Nov 01 13:00:52 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.596	0.612	-2.7	106	0.00
3	n-Nitrosodimethylamine	0.797	0.844	-5.9	117	0.00
4 S	2-Fluorophenol	1.187	1.125	5.2	95	0.00
5 S	Phenol-d6	1.955	1.877	4.0	99	0.00
6	bis(2-Chloroethyl)ether	1.431	1.356	5.2	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.01
8 S	Nitrobenzene-d5	0.429	0.419	2.3	103	0.00
9	Naphthalene	3.998	3.865	3.3	101	0.00
10	Hexachlorobutadiene	0.262	0.262	0.0	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.705	0.681	3.4	99	0.00
12	2-Methylnaphthalene	0.714	0.681	4.6	101	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.189	0.160	15.3	98	0.00
15 S	2-Fluorobiphenyl	1.645	1.533	6.8	100	0.00
16	Acenaphthylene	1.825	1.723	5.6	99	0.00
17	Acenaphthene	1.096	1.078	1.6	100	0.00
18	Fluorene	1.454	1.380	5.1	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4-Bromophenyl-phenylether	0.244	0.228	6.6	98	0.00
21	Hexachlorobenzene	0.255	0.242	5.1	101	0.00
22	Pentachlorophenol	0.057	0.040	29.8#	82	0.00
23	Phenanthrene	0.991	0.922	7.0	99	0.00
24	Anthracene	0.942	0.884	6.2	101	0.00
25 SURR	Fluoranthene-d10	5.159	4.707	8.8	98	0.00
26	Fluoranthene	1.241	1.146	7.7	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pyrene	1.092	1.015	7.1	100	0.00
29 S	Terphenyl-d14	0.930	0.842	9.5	98	0.00
30	Benzo(a)anthracene	1.169	1.077	7.9	99	0.00
31	Chrysene	1.156	1.070	7.4	99	0.01
32	Bis(2-ethylhexyl)phthalate	4.140	3.491	15.7	104	0.01
33	Indeno(1,2,3-cd)pyrene	1.203	1.127	6.3	100	0.00
34 I	Perylene-d12	1.000	1.000	0.0	99	0.00
35	Benzo(b)fluoranthene	1.127	1.062	5.8	100	0.00
36	Benzo(k)fluoranthene	1.166	1.098	5.8	101	0.00
37 C	Benzo(a)pyrene	1.101	1.021	7.3	98	0.00
38	Dibenzo(a,h)anthracene	1.027	0.953	7.2	97	0.00
39	Benzo(a,h,i)perylene	1.022	0.946	7.4	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE110118\  
Data File : BE098143.D  
Acq On : 1 Nov 2018 13:04  
Operator : SJ/JU  
Sample : SSTDICV0.4  
Misc :  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
ICVBE110118

Quant Time: Nov 01 13:34:12 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE110118.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Nov 01 13:00:52 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
-----				
(#) = Out of Range	SPCC's out = 0		CCC's out = 0	