

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE110118\
 Data File : BE098145.D
 Acq On : 1 Nov 2018 14:18
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 01 14:51:52 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE110118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 13:00:52 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.596	0.644	-8.1	90	0.00
3	n-Nitrosodimethylamine	0.797	0.838	-5.1	93	0.00
4 S	2-Fluorophenol	1.187	1.303	-9.8	89	0.00
5 S	Phenol-d6	1.955	2.101	-7.5	89	0.00
6	bis(2-Chloroethyl)ether	1.431	1.452	-1.5	87	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
8 S	Nitrobenzene-d5	0.429	0.424	1.2	88	0.00
9	Naphthalene	3.998	4.125	-3.2	91	0.00
10	Hexachlorobutadiene	0.262	0.276	-5.3	93	0.00
11 SURR	2-Methylnaphthalene-d10	0.705	0.729	-3.4	89	0.00
12	2-Methylnaphthalene	0.714	0.742	-3.9	93	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.189	0.171	9.5	90	0.00
15 S	2-Fluorobiphenyl	1.645	1.671	-1.6	94	0.00
16	Acenaphthylene	1.825	1.859	-1.9	92	0.00
17	Acenaphthene	1.096	1.132	-3.3	90	0.00
18	Fluorene	1.454	1.450	0.3	89	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
20	4-Bromophenyl-phenylether	0.244	0.252	-3.3	92	0.00
21	Hexachlorobenzene	0.255	0.265	-3.9	94	0.00
22	Pentachlorophenol	0.057	0.046	19.3	80	0.00
23	Phenanthrene	0.991	1.009	-1.8	92	0.00
24	Anthracene	0.942	0.963	-2.2	94	0.00
25 SURR	Fluoranthene-d10	5.159	5.170	-0.2	92	0.00
26	Fluoranthene	1.241	1.242	-0.1	92	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
28	Pvrene	1.092	1.147	-5.0	92	0.00
29 S	Terphenyl-d14	0.930	0.966	-3.9	92	0.00
30	Benzo(a)anthracene	1.169	1.213	-3.8	91	0.00
31	Chrysene	1.156	1.200	-3.8	91	0.00
32	Bis(2-ethylhexyl)phthalate	4.140	4.086	1.3	100	0.01
33	Indeno(1,2,3-cd)pyrene	1.203	1.237	-2.8	90	0.00
34 I	Perylene-d12	1.000	1.000	0.0	82	0.00
35	Benzo(b)fluoranthene	1.127	1.151	-2.1	89	0.00
36	Benzo(k)fluoranthene	1.166	1.225	-5.1	92	0.00
37 C	Benzo(a)pyrene	1.101	1.134	-3.0	89	0.00
38	Dibenzo(a,h)anthracene	1.027	1.061	-3.3	89	0.00
39	Benzo(a,h,i)perylene	1.022	1.063	-4.0	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE110118\
Data File : BE098145.D
Acq On : 1 Nov 2018 14:18
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 01 14:51:52 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE110118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 13:00:52 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			