

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE110518\
 Data File : BE098161.D
 Acq On : 5 Nov 2018 14:42
 Operator : SJ/JU
 Sample : J5780-05
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 CPS113081

Quant Time: Nov 06 10:26:42 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 13:00:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.86	152	945	0.40	ng	-0.01
7) Naphthalene-d8	10.67	136	4467	0.40	ng	0.00
13) Acenaphthene-d10	14.53	164	3230	0.40	ng	-0.01
19) Phenanthrene-d10	17.28	188	7228	0.40	ng	0.00
27) Chrysene-d12	21.46	240	8219	0.40	ng	0.00
34) Perylene-d12	24.01	264	7882	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.46	112	520	0.19	ng	0.00
5) Phenol-d6	7.05	99	507	0.11	ng	0.00
8) Nitrobenzene-d5	9.02	82	1713	0.36	ng	-0.01
11) 2-Methylnaphthalene-d10	12.27	152	3053	0.39	ng	0.00
14) 2,4,6-Tribromophenol	16.02	330	615	0.40	ng	-0.01
15) 2-Fluorobiphenyl	13.15	172	5958	0.45	ng	-0.01
25) Fluoranthene-d10	19.31	212	35841	0.38	ng	0.00
29) Terphenyl-d14	19.91	244	8544	0.45	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Naphthalene	10.72	128	3702	0.083	ng	# 62
22) Pentachlorophenol	16.94	266	59	0.177	ng	# 71
26) Fluoranthene	19.34	202	1368	0.061	ng	# 97
28) Pyrene	19.70	202	2872	0.128	ng	98
32) Bis(2-ethylhexyl)phthalate	21.37	149	13685	0.106	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE110518\
Data File : BE098161.D
Acq On : 5 Nov 2018 14:42
Operator : SJ/JU
Sample : J5780-05
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
CPS113081

Quant Time: Nov 06 10:26:42 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 13:00:52 2018
Response via : Initial Calibration

