

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE110518\
 Data File : BE098153.D
 Acq On : 5 Nov 2018 9:41
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 06 10:21:34 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 13:00:52 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	81	-0.01
2	1,4-Dioxane	0.400	0.386	3.5	78	0.00
3	n-Nitrosodimethylamine	0.400	0.394	1.5	85	0.00
4 S	2-Fluorophenol	0.400	0.382	4.5	76	0.00
5 S	Phenol-d6	0.400	0.352	12.0	71	0.00
6	bis(2-Chloroethyl)ether	0.400	0.393	1.8	82	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	78	0.00
8 S	Nitrobenzene-d5	0.400	0.354	11.5	71	0.00
9	Naphthalene	0.400	0.421	-5.2	84	0.00
10	Hexachlorobutadiene	0.400	0.418	-4.5	84	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.399	0.3	78	0.00
12	2-Methylnaphthalene	0.400	0.416	-4.0	83	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	72	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.351	12.3	75	0.00
15 S	2-Fluorobiphenyl	0.400	0.412	-3.0	82	0.00
16	Acenaphthylene	0.400	0.415	-3.7	80	0.00
17	Acenaphthene	0.400	0.421	-5.2	79	0.00
18	Fluorene	0.400	0.411	-2.7	80	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	76	0.00
20	4-Bromophenyl-phenylether	0.400	0.418	-4.5	82	0.00
21	Hexachlorobenzene	0.400	0.426	-6.5	85	0.00
22	Pentachlorophenol	0.400	0.349	12.8	62	0.00
23	Phenanthrene	0.400	0.417	-4.2	83	0.00
24	Anthracene	0.400	0.407	-1.7	83	-0.01
25 SURR	Fluoranthene-d10	0.400	0.418	-4.5	85	0.00
26	Fluoranthene	0.400	0.424	-6.0	86	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	78	0.00
28	Pyrene	0.400	0.411	-2.7	87	0.00
29 S	Terphenyl-d14	0.400	0.412	-3.0	88	0.00
30	Benzo(a)anthracene	0.400	0.407	-1.7	86	0.00
31	Chrysene	0.400	0.417	-4.2	88	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.132	67.0#	43	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.398	0.5	84	0.00
34 I	Perylene-d12	0.400	0.400	0.0	80	0.00
35	Benzo(b)fluoranthene	0.400	0.406	-1.5	86	0.00
36	Benzo(k)fluoranthene	0.400	0.397	0.8	85	0.00
37 C	Benzo(a)pyrene	0.400	0.394	1.5	83	0.00
38	Dibenzo(a,h)anthracene	0.400	0.398	0.5	84	0.00
39	Benzo(g,h,i)perylene	0.400	0.396	1.0	84	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE110518\
Data File : BE098153.D
Acq On : 5 Nov 2018 9:41
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 06 10:21:34 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 13:00:52 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				