

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE110718\
 Data File : BE098177.D
 Acq On : 7 Nov 2018 17:04
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 08 08:46:47 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 13:00:52 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	-0.01
2	1,4-Dioxane	0.596	0.638	-7.0	97	0.00
3	n-Nitrosodimethylamine	0.797	0.816	-2.4	99	0.00
4 S	2-Fluorophenol	1.187	1.230	-3.6	92	0.00
5 S	Phenol-d6	1.955	2.087	-6.8	97	0.00
6	bis(2-Chloroethyl)ether	1.431	1.468	-2.6	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
8 S	Nitrobenzene-d5	0.429	0.424	1.2	95	-0.01
9	Naphthalene	3.998	4.044	-1.2	96	0.00
10	Hexachlorobutadiene	0.262	0.263	-0.4	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.705	0.717	-1.7	94	0.00
12	2-Methylnaphthalene	0.714	0.736	-3.1	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	85	-0.01
14 S	2,4,6-Tribromophenol	0.189	0.192	-1.6	102	0.00
15 S	2-Fluorobiphenyl	1.645	1.703	-3.5	97	-0.01
16	Acenaphthylene	1.825	1.973	-8.1	99	0.00
17	Acenaphthene	1.096	1.213	-10.7	98	0.00
18	Fluorene	1.454	1.572	-8.1	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
20	4-Bromophenyl-phenylether	0.244	0.259	-6.1	100	-0.01
21	Hexachlorobenzene	0.255	0.267	-4.7	100	0.00
22	Pentachlorophenol	0.057	0.069	-21.1	125	0.00
23	Phenanthrene	0.991	1.011	-2.0	97	0.00
24	Anthracene	0.942	0.963	-2.2	99	-0.01
25 SURR	Fluoranthene-d10	5.159	5.231	-1.4	98	0.00
26	Fluoranthene	1.241	1.270	-2.3	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
28	Pyrene	1.092	1.144	-4.8	98	0.00
29 S	Terphenyl-d14	0.930	0.969	-4.2	98	0.00
30	Benzo(a)anthracene	1.169	1.219	-4.3	97	0.00
31	Chrysene	1.156	1.217	-5.3	98	0.00
32	Bis(2-ethylhexyl)phthalate	4.140	2.875	30.6#	75	0.00
33	Indeno(1,2,3-cd)pyrene	1.203	1.224	-1.7	95	0.00
34 I	Perylene-d12	1.000	1.000	0.0	87	0.00
35	Benzo(b)fluoranthene	1.127	1.142	-1.3	93	0.00
36	Benzo(k)fluoranthene	1.166	1.226	-5.1	98	0.00
37 C	Benzo(a)pyrene	1.101	1.126	-2.3	94	0.00
38	Dibenzo(a,h)anthracene	1.027	1.036	-0.9	92	0.00
39	Benzo(g,h,i)perylene	1.022	1.034	-1.2	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE110718\
Data File : BE098177.D
Acq On : 7 Nov 2018 17:04
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 08 08:46:47 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 13:00:52 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			