

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE110719\
 Data File : BE100700.D
 Acq On : 7 Nov 2019 13:57
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 08 11:18:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 16:49:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	0.00
2	1,4-Dioxane	0.714	0.645	9.7	65	0.00
3	n-Nitrosodimethylamine	0.755	0.680	9.9	70	0.00
4 S	2-Fluorophenol	1.235	1.220	1.2	75	0.00
5 S	Phenol-d6	1.742	1.906	-9.4	83	0.00
6	bis(2-Chloroethyl)ether	1.430	1.413	1.2	75	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	75	0.00
8 S	Nitrobenzene-d5	0.248	0.320	-29.0#	111	0.00
9	Naphthalene	4.206	4.143	1.5	76	0.00
10	Hexachlorobutadiene	0.183	0.182	0.5	78	0.00
11 SURR	2-Methylnaphthalene-d10	0.631	0.610	3.3	76	0.00
12	2-Methylnaphthalene	0.717	0.701	2.2	77	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	81	0.00
14 S	2,4,6-Tribromophenol	0.106	0.134	-26.4#	117	0.00
15 S	2-Fluorobiphenyl	1.457	1.545	-6.0	88	0.00
16	Acenaphthylene	1.721	1.454	15.5	71	-0.01
17	Acenaphthene	1.112	1.054	5.2	80	0.00
18	Fluorene	1.477	1.420	3.9	81	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	91	-0.01
20	4-Bromophenyl-phenylether	0.204	0.181	11.3	84	0.00
21	Hexachlorobenzene	0.194	0.184	5.2	91	-0.01
22	Pentachlorophenol	0.062	0.047	24.2	77	0.01
23	Phenanthrene	1.100	1.106	-0.5	91	-0.01
24	Anthracene	1.021	0.866	15.2	80	0.00
25 SURR	Fluoranthene-d10	5.162	4.795	7.1	88	0.00
26	Fluoranthene	1.436	1.407	2.0	89	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
28	Pvrene	1.118	1.205	-7.8	88	0.00
29 S	Terphenyl-d14	0.875	1.017	-16.2	100	0.00
30	Benzo(a)anthracene	1.305	1.196	8.4	75	0.00
31	Chrysene	1.304	1.238	5.1	78	0.00
32	Bis(2-ethylhexyl)phthalate	2.729	1.854	32.1#	55	-0.01
33	Indeno(1,2,3-cd)pyrene	1.603	1.359	15.2	71	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	74	0.00
35	Benzo(b)fluoranthene	1.166	1.047	10.2	72	0.00
36	Benzo(k)fluoranthene	1.200	1.112	7.3	74	0.00
37 C	Benzo(a)pyrene	1.086	0.984	9.4	73	-0.01
38	Dibenzo(a,h)anthracene	1.122	0.969	13.6	71	-0.01
39	Benzo(a,h,i)perylene	1.129	1.044	7.5	74	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE110719\
Data File : BE100700.D
Acq On : 7 Nov 2019 13:57
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 08 11:18:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 16:49:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			