

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE110818\
 Data File : BE098180.D
 Acq On : 8 Nov 2018 10:37
 Operator : SJ/JU
 Sample : J5820-03DL 50X
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 CPS093081DL

Quant Time: Nov 09 09:51:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 13:00:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.87	152	1215	0.40	ng	0.00
7) Naphthalene-d8	10.68	136	5916	0.40	ng	0.01
13) Acenaphthene-d10	14.54	164	4156	0.40	ng	0.00
19) Phenanthrene-d10	17.28	188	10164	0.40	ng	0.00
27) Chrysene-d12	21.46	240	11266	0.40	ng	0.00
34) Perylene-d12	24.02	264	11003	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.47	112	28	0.01	ng	0.00
5) Phenol-d6	7.07	99	20	0.00	ng	0.02
8) Nitrobenzene-d5	9.06	82	65	0.01	ng	0.03
11) 2-Methylnaphthalene-d10	12.28	152	67	0.01	ng	0.02
14) 2,4,6-Tribromophenol	16.05	330	24	0.01	ng	0.01
15) 2-Fluorobiphenyl	13.16	172	128	0.01	ng	0.00
25) Fluoranthene-d10	19.32	212	796	0.01	ng	0.00
29) Terphenyl-d14	19.91	244	243	0.01	ng	0.00
Target Compounds						Qvalue
22) Pentachlorophenol	16.94	266	1638	0.989	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE110818\
Data File : BE098180.D
Acq On : 8 Nov 2018 10:37
Operator : SJ/JU
Sample : J5820-03DL 50X
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
CPS093081DL

Quant Time: Nov 09 09:51:24 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 13:00:52 2018
Response via : Initial Calibration

