

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE110818\
 Data File : BE098179.D
 Acq On : 8 Nov 2018 9:28
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Nov 09 09:48:11 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 13:00:52 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.596	0.610	-2.3	102	0.00
3	n-Nitrosodimethylamine	0.797	0.787	1.3	105	0.00
4 S	2-Fluorophenol	1.187	1.183	0.3	97	0.00
5 S	Phenol-d6	1.955	1.908	2.4	97	0.00
6	bis(2-Chloroethyl)ether	1.431	1.434	-0.2	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.429	0.421	1.9	101	-0.01
9	Naphthalene	3.998	4.047	-1.2	103	0.00
10	Hexachlorobutadiene	0.262	0.257	1.9	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.705	0.705	0.0	100	0.00
12	2-Methylnaphthalene	0.714	0.729	-2.1	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	-0.01
14 S	2,4,6-Tribromophenol	0.189	0.156	17.5	91	0.00
15 S	2-Fluorobiphenyl	1.645	1.644	0.1	102	-0.01
16	Acenaphthylene	1.825	1.853	-1.5	102	0.00
17	Acenaphthene	1.096	1.119	-2.1	99	0.00
18	Fluorene	1.454	1.475	-1.4	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
20	4-Bromophenyl-phenylether	0.244	0.250	-2.5	98	0.00
21	Hexachlorobenzene	0.255	0.269	-5.5	103	0.00
22	Pentachlorophenol	0.057	0.054	5.3	100	0.00
23	Phenanthrene	0.991	1.035	-4.4	101	0.00
24	Anthracene	0.942	0.954	-1.3	100	-0.01
25 SURR	Fluoranthene-d10	5.159	5.194	-0.7	99	0.00
26	Fluoranthene	1.241	1.278	-3.0	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
28	Pyrene	1.092	1.168	-7.0	103	0.00
29 S	Terphenyl-d14	0.930	0.964	-3.7	100	0.00
30	Benzo(a)anthracene	1.169	1.207	-3.3	99	0.00
31	Chrysene	1.156	1.202	-4.0	100	0.00
32	Bis(2-ethylhexyl)phthalate	4.140	2.653	35.9#	71	0.00
33	Indeno(1,2,3-cd)pyrene	1.203	1.162	3.4	93	0.00
34 I	Perylene-d12	1.000	1.000	0.0	88	0.00
35	Benzo(b)fluoranthene	1.127	1.185	-5.1	98	0.00
36	Benzo(k)fluoranthene	1.166	1.224	-5.0	99	0.00
37 C	Benzo(a)pyrene	1.101	1.128	-2.5	96	0.00
38	Dibenzo(a,h)anthracene	1.027	1.020	0.7	92	0.00
39	Benzo(g,h,i)perylene	1.022	1.019	0.3	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE110818\
Data File : BE098179.D
Acq On : 8 Nov 2018 9:28
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 09 09:48:11 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 13:00:52 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			