

Data Path : Z:\svoasrv\HPCHEM1\BNA\_E\Data\BE110818\  
 Data File : BE098181.D  
 Acq On : 8 Nov 2018 13:06  
 Operator : SJ/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Nov 09 09:52:29 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE110118.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Nov 01 13:00:52 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | Amount | Calc. | %Dev  | Area% | Dev(min) |
|---------|----------------------------|--------|-------|-------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 0.400  | 0.400 | 0.0   | 107   | 0.00     |
| 2       | 1,4-Dioxane                | 0.400  | 0.455 | -13.7 | 122   | 0.00     |
| 3       | n-Nitrosodimethylamine     | 0.400  | 0.404 | -1.0  | 116   | 0.00     |
| 4 S     | 2-Fluorophenol             | 0.400  | 0.394 | 1.5   | 103   | 0.00     |
| 5 S     | Phenol-d6                  | 0.400  | 0.394 | 1.5   | 106   | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 0.400  | 0.414 | -3.5  | 115   | 0.00     |
| 7 I     | Naphthalene-d8             | 0.400  | 0.400 | 0.0   | 103   | 0.01     |
| 8 S     | Nitrobenzene-d5            | 0.400  | 0.381 | 4.8   | 101   | 0.00     |
| 9       | Naphthalene                | 0.400  | 0.419 | -4.7  | 110   | 0.01     |
| 10      | Hexachlorobutadiene        | 0.400  | 0.396 | 1.0   | 104   | 0.01     |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.400  | 0.409 | -2.2  | 105   | 0.02     |
| 12      | 2-Methylnaphthalene        | 0.400  | 0.425 | -6.2  | 112   | 0.02     |
| 13 I    | Acenaphthene-d10           | 0.400  | 0.400 | 0.0   | 97    | 0.00     |
| 14 S    | 2,4,6-Tribromophenol       | 0.400  | 0.344 | 14.0  | 99    | 0.00     |
| 15 S    | 2-Fluorobiphenyl           | 0.400  | 0.404 | -1.0  | 108   | 0.00     |
| 16      | Acenaphthylene             | 0.400  | 0.410 | -2.5  | 107   | 0.00     |
| 17      | Acenaphthene               | 0.400  | 0.415 | -3.7  | 105   | 0.00     |
| 18      | Fluorene                   | 0.400  | 0.405 | -1.3  | 105   | 0.00     |
| 19 I    | Phenanthrene-d10           | 0.400  | 0.400 | 0.0   | 97    | 0.01     |
| 20      | 4-Bromophenyl-phenylether  | 0.400  | 0.409 | -2.2  | 103   | 0.00     |
| 21      | Hexachlorobenzene          | 0.400  | 0.409 | -2.2  | 104   | 0.00     |
| 22      | Pentachlorophenol          | 0.400  | 0.445 | -11.2 | 114   | 0.01     |
| 23      | Phenanthrene               | 0.400  | 0.419 | -4.7  | 106   | 0.00     |
| 24      | Anthracene                 | 0.400  | 0.409 | -2.2  | 106   | 0.00     |
| 25 SURR | Fluoranthene-d10           | 0.400  | 0.390 | 2.5   | 100   | 0.00     |
| 26      | Fluoranthene               | 0.400  | 0.400 | 0.0   | 104   | 0.00     |
| 27 I    | Chrysene-d12               | 0.400  | 0.400 | 0.0   | 90    | 0.00     |
| 28      | Pyrene                     | 0.400  | 0.419 | -4.7  | 103   | 0.00     |
| 29 S    | Terphenyl-d14              | 0.400  | 0.411 | -2.7  | 102   | 0.00     |
| 30      | Benzo(a)anthracene         | 0.400  | 0.399 | 0.3   | 98    | 0.00     |
| 31      | Chrysene                   | 0.400  | 0.414 | -3.5  | 101   | 0.01     |
| 32      | Bis(2-ethylhexyl)phthalate | 0.400  | 0.248 | 38.0# | 70    | 0.01     |
| 33      | Indeno(1,2,3-cd)pyrene     | 0.400  | 0.388 | 3.0   | 95    | 0.00     |
| 34 I    | Perylene-d12               | 0.400  | 0.400 | 0.0   | 91    | 0.00     |
| 35      | Benzo(b)fluoranthene       | 0.400  | 0.415 | -3.7  | 100   | 0.00     |
| 36      | Benzo(k)fluoranthene       | 0.400  | 0.424 | -6.0  | 103   | 0.00     |
| 37 C    | Benzo(a)pyrene             | 0.400  | 0.408 | -2.0  | 98    | 0.00     |
| 38      | Dibenzo(a,h)anthracene     | 0.400  | 0.393 | 1.8   | 94    | 0.00     |
| 39      | Benzo(g,h,i)perylene       | 0.400  | 0.403 | -0.8  | 97    | 0.01     |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_E\Data\BE110818\  
Data File : BE098181.D  
Acq On : 8 Nov 2018 13:06  
Operator : SJ/JU  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_E  
LabSampleId :  
SSTDCCC0.4

Quant Time: Nov 09 09:52:29 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE110118.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Nov 01 13:00:52 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound           | Amount                       | Calc. | %Dev | Area% | Dev(min) |
|--------------------|------------------------------|-------|------|-------|----------|
| -----              |                              |       |      |       |          |
| (#) = Out of Range | SPCC's out = 0 CCC's out = 0 |       |      |       |          |