

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE110818\
 Data File : BE098181.D
 Acq On : 8 Nov 2018 13:06
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 09 09:52:29 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 13:00:52 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.596	0.678	-13.8	122	0.00
3	n-Nitrosodimethylamine	0.797	0.805	-1.0	116	0.00
4 S	2-Fluorophenol	1.187	1.170	1.4	103	0.00
5 S	Phenol-d6	1.955	1.928	1.4	106	0.00
6	bis(2-Chloroethyl)ether	1.431	1.481	-3.5	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.01
8 S	Nitrobenzene-d5	0.429	0.409	4.7	101	0.00
9	Naphthalene	3.998	4.192	-4.9	110	0.01
10	Hexachlorobutadiene	0.262	0.259	1.1	104	0.01
11 SURR	2-Methylnaphthalene-d10	0.705	0.721	-2.3	105	0.02
12	2-Methylnaphthalene	0.714	0.758	-6.2	112	0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.189	0.163	13.8	99	0.00
15 S	2-Fluorobiphenyl	1.645	1.661	-1.0	108	0.00
16	Acenaphthylene	1.825	1.872	-2.6	107	0.00
17	Acenaphthene	1.096	1.137	-3.7	105	0.00
18	Fluorene	1.454	1.473	-1.3	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.01
20	4-Bromophenyl-phenylether	0.244	0.250	-2.5	103	0.00
21	Hexachlorobenzene	0.255	0.261	-2.4	104	0.00
22	Pentachlorophenol	0.057	0.059	-3.5	114	0.01
23	Phenanthrene	0.991	1.037	-4.6	106	0.00
24	Anthracene	0.942	0.962	-2.1	106	0.00
25 SURR	Fluoranthene-d10	5.159	5.026	2.6	100	0.00
26	Fluoranthene	1.241	1.242	-0.1	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
28	Pyrene	1.092	1.144	-4.8	103	0.00
29 S	Terphenyl-d14	0.930	0.955	-2.7	102	0.00
30	Benzo(a)anthracene	1.169	1.165	0.3	98	0.00
31	Chrysene	1.156	1.197	-3.5	101	0.01
32	Bis(2-ethylhexyl)phthalate	4.140	2.576	37.8#	70	0.01
33	Indeno(1,2,3-cd)pyrene	1.203	1.166	3.1	95	0.00
34 I	Perylene-d12	1.000	1.000	0.0	91	0.00
35	Benzo(b)fluoranthene	1.127	1.169	-3.7	100	0.00
36	Benzo(k)fluoranthene	1.166	1.236	-6.0	103	0.00
37 C	Benzo(a)pyrene	1.101	1.123	-2.0	98	0.00
38	Dibenzo(a,h)anthracene	1.027	1.009	1.8	94	0.00
39	Benzo(g,h,i)perylene	1.022	1.029	-0.7	97	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE110818\
Data File : BE098181.D
Acq On : 8 Nov 2018 13:06
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 09 09:52:29 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE110118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 13:00:52 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			