

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE110819\
 Data File : BE100711.D
 Acq On : 8 Nov 2019 10:49
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 08 12:40:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 16:49:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	41	0.01
2	1,4-Dioxane	0.400	0.322	19.5	33	0.00
3	n-Nitrosodimethylamine	0.400	0.356	11.0	39	0.00
4 S	2-Fluorophenol	0.400	0.442	-10.5	47	0.01
5 S	Phenol-d6	0.400	0.486	-21.5	52	0.02
6	bis(2-Chloroethyl)ether	0.400	0.384	4.0	42	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	46	0.01
8 S	Nitrobenzene-d5	0.400	0.583	-45.7#	77	0.01
9	Naphthalene	0.400	0.385	3.8	46	0.01
10	Hexachlorobutadiene	0.400	0.401	-0.3	48	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.403	-0.8	49	0.01
12	2-Methylnaphthalene	0.400	0.397	0.8	48	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	50	0.01
14 S	2,4,6-Tribromophenol	0.400	0.174	56.5#	24	0.03
15 S	2-Fluorobiphenyl	0.400	0.408	-2.0	52	0.01
16	Acenaphthylene	0.400	0.380	5.0	49	0.00
17	Acenaphthene	0.400	0.367	8.3	47	0.01
18	Fluorene	0.400	0.377	5.8	48	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	48	0.00
20	4-Bromophenyl-phenylether	0.400	0.403	-0.8	50	0.00
21	Hexachlorobenzene	0.400	0.403	-0.8	51	0.00
22	Pentachlorophenol	0.400	0.004	99.0#	1	0.00
23	Phenanthrene	0.400	0.397	0.8	47	0.00
24	Anthracene	0.400	0.395	1.3	49	0.01
25 SURR	Fluoranthene-d10	0.400	0.377	5.8	47	0.01
26	Fluoranthene	0.400	0.385	3.8	46	0.01
27 I	Chrysene-d12	0.400	0.400	0.0	41	0.01
28	Pvrene	0.400	0.445	-11.2	46	0.00
29 S	Terphenyl-d14	0.400	0.474	-18.5	52	0.00
30	Benzo(a)anthracene	0.400	0.399	0.3	42	0.01
31	Chrysene	0.400	0.385	3.8	40	0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.484	-21.0	50	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.379	5.3	41	0.04
34 I	Perylene-d12	0.400	0.400	0.0	38	0.02
35	Benzo(b)fluoranthene	0.400	0.385	3.8	39	0.02
36	Benzo(k)fluoranthene	0.400	0.379	5.3	38	0.01
37 C	Benzo(a)pyrene	0.400	0.402	-0.5	41	0.02
38	Dibenzo(a,h)anthracene	0.400	0.394	1.5	41	0.04
39	Benzo(a,h,i)perylene	0.400	0.384	4.0	39	0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE110819\
Data File : BE100711.D
Acq On : 8 Nov 2019 10:49
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 08 12:40:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 16:49:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				